



Alcohol Testing Program

**INSTRUMENT PROCESSING SHEET**

Agency Bay County S/N 80-001309  
Date In 6/17/17 Date Out 6/19/2017  Ship  P/U  H/D  CMI  EE

**Intake** Performed By AS

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By OSM

Lab Temp °C 21.6  
 Breath Tube Screen  
 Replace External O-Rings  
 Instrument Set Up Verified  
 R-Value 198  
 Flow Verification (L/s)  
 Flow Column # ATP100  
 32mm .148 (.139 - .169)  
 36mm .167 (.156 - .190)  
 53mm .234 (.228 - .278)  
 103mm .492 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G11739</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>G8149</u>	<u>201611B</u> <u>11/15/18</u>
0.20	<u>G11621</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG626604</u> <u>9/22/18</u>

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By OSM

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other Loaded forms

**Quality Checks Cont.** Performed By OSM

**Simulator Temperatures °C**  
 External Digital Therm. ID#: 300505  
 34°C +/- .2 Serial #: G11739  
 34°C +/- .2 Serial #: G8149  
 34°C +/- .2 Serial #: G11621

**Calibration Adjustment** Performed By \_\_\_\_\_

Calibration Adjustment N/A  
 Calibration Adjustment Complete  
 Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By OSM

Barometric Pressure 1013 Gauge  
 ID# 26932 1011 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>G2879</u>
Interferent	<u>G8144</u>
0.05	<u>G11739</u>
0.08	<u>G8149</u>
0.20	<u>G11621</u>

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Calibration Adjustment  
 Post-Stability Tests  
 Other forms 38, 40, 41

Notes/Suggested Service:  
Enforcement Electronics did NOT load forms.  
I loaded AND tested forms

OSM 6/20/17 OSM 6/20/17

Quality Control Review \_\_\_\_\_ Date \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001309  
06/19/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:15
Control Test	0.049	08:15
Air Blank	0.000	08:16
Control Test	0.049	08:16
Air Blank	0.000	08:17
Control Test	0.050	08:18
Air Blank	0.000	08:18
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

*P. Murphy*  
-----  
Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001309  
06/19/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:25
Control Test	0.080	08:26
Air Blank	0.000	08:27
Control Test	0.080	08:27
Air Blank	0.000	08:28
Control Test	0.080	08:29
Air Blank	0.000	08:29
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*P. Murphy*  
-----  
Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001309  
06/19/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:20
Control Test	0.194	08:21
Air Blank	0.000	08:21
Control Test	0.195	08:22
Air Blank	0.000	08:22
Control Test	0.197	08:23
Air Blank	0.000	08:24
Control Test Stats		
Average	0.1953	
Std Dev	0.0015	
Rel Std Dev(%)	0.7820	

*P. Murphy*  
-----  
Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001309  
06/19/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.078	08:36
Air Blank	0.000	08:36
Control Test	0.079	08:37
Air Blank	0.000	08:37
Control Test	0.078	08:37
Air Blank	0.000	08:38
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS

*P. Murphy*  
-----  
Operator's Signature

*ja*