



INSTRUMENT PROCESSING SHEET

Agency Walton County

S/N 80-001312

Florida Department of Law Enforcement

Date In 10/31/17

DI Completion Date 11/3/17

Ship P/U H/D CMI EE

Intake Performed By [Signature]

Annual
 Registration
 Return from CMI / EE

Visual Inspection:
 Case Handle
 Keyboard Dry Gas Shelf
 Feet Breath Tube
 Ports Screws Tight

Other Equipment/ Accessories:
 Power cord Printer Cable
 Static Bag 12V DC Cable

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 120
 Flow Verification (L/s)
 Flow Column # ATP105
 32 mm .156 (.139 - .169)
 36 mm .171 (.156 - .190)
 53 mm .242 (.228 - .278)
 103 mm .500 (.447 - .547)
 Barometric Pressure Check
 Gauge ID # 28421
 Stability Checks

Flow Calibration Performed By _____

Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32 mm _____ (.139 - .169)
 36 mm _____ (.156 - .190)
 53 mm _____ (.228 - .278)
 103 mm _____ (.447 - .547)

Final Release Date

NOV 06 2017

Simulator	Serial #	Lot #/Exp
0.050	<u>G2835</u>	201707D 07/25/2019
0.080	<u>DR1279</u>	201707E 07/25/2019
0.200	<u>SD1025</u>	201707C 07/24/2019
0.080 DGS	N/A	<u>A6626605</u> 9/22/18

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By [Signature]

Lab Temp °C 22.3
 External Digital Therm. ID#: 300504
 34°C +/- .2 Serial #: G2835
 34°C +/- .2 Serial #: DR1279
 34°C +/- .2 Serial #: SD1025

Calibration Adjustment Performed By _____

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure ID# 28421
 Gauge 1019 Instrument 1019
 Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	<u>G2880</u>
Interferent	<u>G2840</u>
0.050	<u>G2835</u>
0.080	<u>DR1279</u>
0.200	<u>SD1025</u>

Attachments

Form 41 Flow Calibration
 Stability Checks Form 40
 Calibration Adjustment Other _____
 Post-Stability Checks

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature] 11/6/17 [Signature] 11/6/17
 Tech Review / Date Admin Review / Date

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001312
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:13
Control Test	0.053	12:14
Air Blank	0.000	12:15
Control Test	0.052	12:15
Air Blank	0.000	12:16
Control Test	0.051	12:17
Air Blank	0.000	12:17
Control Test Stats		
Average	0.0520	
Std Dev	0.0010	
Rel Std Dev(%)	1.9231	


Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001312
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:29
Control Test	0.083	12:30
Air Blank	0.000	12:30
Control Test	0.082	12:31
Air Blank	0.000	12:32
Control Test	0.082	12:32
Air Blank	0.000	12:33
Control Test Stats		
Average	0.0823	
Std Dev	0.0006	
Rel Std Dev(%)	0.7012	


Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001312
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:24
Control Test	0.199	12:25
Air Blank	0.000	12:25
Control Test	0.200	12:26
Air Blank	0.000	12:26
Control Test	0.202	12:27
Air Blank	0.000	12:28
Control Test Stats		
Average	0.2003	
Std Dev	0.0015	
Rel Std Dev(%)	0.7625	


Operator's Signature

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001312
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:44
Control Test	0.080	12:45
Air Blank	0.000	12:45
Control Test	0.079	12:45
Air Blank	0.000	12:46
Control Test	0.080	12:46
Air Blank	0.000	12:47
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS


Operator's Signature



11/6/17
JD