

Agency Okeechobee County S/N 80-001321 MAY 01 2017
 Date In 4/20/17 Date Out 4/26/17 Ship P/U H/D CMI FDLE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 197
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .160 (.139 - .169)
 36mm .179 (.156 - .190)
 53mm .234 (.228 - .278)
 103mm .503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G3709</u>	<u>201603D</u> <u>3-8-8</u>
0.08	<u>DR1279</u>	<u>201611B</u> <u>11-15-18</u>
0.20	<u>DR3856</u>	<u>201604C</u> <u>4-5-18</u>
0.08 DGS	N/A	<u>AG626605</u> <u>9-22-18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure 1009 Gauge
 ID# 28427 1010 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2850</u>
Interferent	<u>G2834</u>
0.05	<u>G3709</u>
0.08	<u>DR1279</u>
0.20	<u>DR3856</u>

Attachments

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: QA/QC OK [Signature] 4/26/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature]
 Quality Control Review

5/1/17
 Date

STABILITY CHECKS - INSTRUMENT # 80-001321 - OKEECHOBEE COUNTY SO

OKEECHOBEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001321
 04/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:01
Control Test	0.048	10:02
Air Blank	0.000	10:03
Control Test	0.050	10:03
Air Blank	0.000	10:04
Control Test	0.050	10:04
Air Blank	0.000	10:05
Control Test Stats		
Average	0.0493	
Std Dev	0.0012	
Rel Std Dev(%)	2.3406	

DFM

SP

Operator's Signature

JD

OKEECHOBEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001321
 04/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:29
Control Test	0.078	10:30
Air Blank	0.000	10:30
Control Test	0.078	10:31
Air Blank	0.000	10:31
Control Test	0.078	10:32
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

OKEECHOBEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001321
 04/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:07
Control Test	0.195	10:07
Air Blank	0.000	10:08
Control Test	0.196	10:09
Air Blank	0.000	10:09
Control Test	0.196	10:10
Air Blank	0.000	10:10
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

SP

Operator's Signature

OKEECHOBEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001321
 04/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:28
Control Test	0.079	09:29
Air Blank	0.000	09:29
Control Test	0.080	09:29
Air Blank	0.000	09:30
Control Test	0.079	09:30
Air Blank	0.000	09:31
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

SP

Operator's Signature