



INSTRUMENT PROCESSING SHEET

Agency Patrick AFBS/N 80-001471

Florida Department of Law Enforcement

Date In 12/22/17DI Completion Date 12/27/17 Ship P/U H/D CMI EE

Intake Performed By <u>SP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>213</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
--	---	--

Final Release Date

FDLE

DEC 27 2017

Alcohol Testing Program

Simulator	Serial #	Lot #/Exp
0.050	SD3962	201707D 7-25-19
0.080	SD1013	201707E 7-25-19
0.200	DR3856	201707C 7-24-19
0.080 DGS	N/A	AG626604 9-22-18

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By SP

Lab Temp °C 21.7
 External Digital Therm. ID#: 300505
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD1013
 34°C +/- .2 Serial #: DR3856

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure ID# 30793
 Gauge 1025 Instrument 1024
 Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	SD1019
Interferent	SD1021
0.050	SD3962
0.080	SD1013
0.200	DR3856

Attachments

Form 41 Flow Calibration
 Stability Checks Form 40
 Calibration Adjustment Other _____
 Post-Stability Checks

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

PJG 12/22/16 JF Pahan 12/27/17
 Tech Review / Date Admin Review / Date

STABILITY CHECKS # 80-001471

PATRICK AFB SFS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001471
12/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:21
Control Test	0.049	07:22
Air Blank	0.000	07:22
Control Test	0.048	07:23
Air Blank	0.000	07:24
Control Test	0.048	07:24
Air Blank	0.000	07:25
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

SP

Operator's Signature

PATRICK AFB SFS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001471
12/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:25
Control Test	0.079	07:26
Air Blank	0.000	07:27
Control Test	0.079	07:27
Air Blank	0.000	07:28
Control Test	0.080	07:29
Air Blank	0.000	07:29
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP

Operator's Signature

PATRICK AFB SFS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001471
12/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:30
Control Test	0.194	07:31
Air Blank	0.000	07:31
Control Test	0.195	07:32
Air Blank	0.000	07:33
Control Test	0.195	07:33
Air Blank	0.000	07:34
Control Test Stats		
Average	0.1947	
Std Dev	0.0006	
Rel Std Dev(%)	0.2966	

SP

Operator's Signature

PATRICK AFB SFS
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001471
12/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:35
Control Test	0.080	07:35
Air Blank	0.000	07:35
Control Test	0.080	07:36
Air Blank	0.000	07:36
Control Test	0.081	07:37
Air Blank	0.000	07:37
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DOS

SP

Operator's Signature

SP

SP
Lylecki