



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Tyndall AFBS/N 80-001646Date In 7/25/17Date Out 9/5/2017☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>SDS</u> ☐ Registration ☒ Annual ☐ Return from CMI ☐ Return from Enforcement Electronics ☐ Other _____ Visual Inspection: ☒ Case ☒ Handle ☒ Dry Gas Holder ☒ Feet ☒ Keyboard/Plug ☒ Back/Plugs ☒ Screws tight ☒ Breath Hose Other Equipment: ☐ Power cord ☐ Printer Cable ☐ Other _____ Notes: _____ _____ _____ _____		Quality Checks Performed By <u>PGM</u> ☒ Lab Temp °C <u>22.5</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>2.15</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.152</u> (.139 - .169) 36mm <u>.167</u> (.156 - .190) 53mm <u>.230</u> (.228 - .278) 103mm <u>.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD 62078</u></td> <td><u>201707D</u></td> </tr> <tr> <td></td> <td><u>SD 62879</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.08</td> <td><u>G8149</u></td> <td><u>201707E</u></td> </tr> <tr> <td></td> <td></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.20</td> <td><u>G11621</u></td> <td><u>201707C</u></td> </tr> <tr> <td></td> <td></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG619605</u></td> </tr> <tr> <td></td> <td></td> <td><u>7/14/18</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	<u>SD 62078</u>	<u>201707D</u>		<u>SD 62879</u>	<u>7/25/19</u>	0.08	<u>G8149</u>	<u>201707E</u>			<u>7/25/19</u>	0.20	<u>G11621</u>	<u>201707C</u>			<u>7/24/19</u>	0.08 DGS	N/A	<u>AG619605</u>			<u>7/14/18</u>	Flow Calibration Performed By _____ ☒ Flow Calibration N/A ☐ Flow Calibration Complete Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)																																
Simulator	Serial #	Lot #/Exp																																																													
0.05	<u>SD 62078</u>	<u>201707D</u>																																																													
	<u>SD 62879</u>	<u>7/25/19</u>																																																													
0.08	<u>G8149</u>	<u>201707E</u>																																																													
		<u>7/25/19</u>																																																													
0.20	<u>G11621</u>	<u>201707C</u>																																																													
		<u>7/24/19</u>																																																													
0.08 DGS	N/A	<u>AG619605</u>																																																													
		<u>7/14/18</u>																																																													
		Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																																													
Calibration Adjustment Performed By <u>PGM</u> ☐ Calibration Adjustment N/A ☒ Calibration Adjustment Complete Barometric Pressure Gauge <u>1013</u> ID # <u>28427</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G6621</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1018</u></td> <td><u>16320</u></td> <td><u>10/21/18</u></td> </tr> <tr> <td>0.100</td> <td><u>SD1012</u></td> <td><u>17060</u></td> <td><u>2/14/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td><u>17090</u></td> <td><u>2/24/19</u></td> </tr> <tr> <td>0.300</td> <td><u>DR1275</u></td> <td><u>16410</u></td> <td><u>12/19/18</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>24515080A3</u></td> <td><u>10/5/17</u></td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3962</u></td> <td><u>201707D</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.08</td> <td><u>SD1013</u></td> <td><u>201707E</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> <td><u>201707C</u></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG619605</u></td> <td><u>7/14/18</u></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>G6621</u>	N/A	N/A	0.040	<u>SD1018</u>	<u>16320</u>	<u>10/21/18</u>	0.100	<u>SD1012</u>	<u>17060</u>	<u>2/14/19</u>	0.200	<u>SD1011</u>	<u>17090</u>	<u>2/24/19</u>	0.300	<u>DR1275</u>	<u>16410</u>	<u>12/19/18</u>	0.080 DGS	N/A	<u>24515080A3</u>	<u>10/5/17</u>	Simulator	Serial Number	Lot Number	Expiration	0.05	<u>SD3962</u>	<u>201707D</u>	<u>7/25/19</u>	0.08	<u>SD1013</u>	<u>201707E</u>	<u>7/25/19</u>	0.20	<u>SD3933</u>	<u>201707C</u>	<u>7/24/19</u>	0.08 DGS	N/A	<u>AG619605</u>	<u>7/14/18</u>	Department Inspection Performed By <u>PGM</u> ☒ Barometric Pressure <u>1014</u> Gauge ID# <u>26932</u> <u>1013</u> Instrument Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>G2879</u></td> </tr> <tr> <td>Interferent</td> <td><u>G8144</u></td> </tr> <tr> <td>0.05</td> <td><u>G2078</u></td> </tr> <tr> <td>0.08</td> <td><u>G8149</u></td> </tr> <tr> <td>0.20</td> <td><u>G11621</u></td> </tr> </tbody> </table>		Simulator	Serial Number	0.00	<u>G2879</u>	Interferent	<u>G8144</u>	0.05	<u>G2078</u>	0.08	<u>G8149</u>	0.20	<u>G11621</u>
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000	<u>G6621</u>	N/A	N/A																																																												
0.040	<u>SD1018</u>	<u>16320</u>	<u>10/21/18</u>																																																												
0.100	<u>SD1012</u>	<u>17060</u>	<u>2/14/19</u>																																																												
0.200	<u>SD1011</u>	<u>17090</u>	<u>2/24/19</u>																																																												
0.300	<u>DR1275</u>	<u>16410</u>	<u>12/19/18</u>																																																												
0.080 DGS	N/A	<u>24515080A3</u>	<u>10/5/17</u>																																																												
Simulator	Serial Number	Lot Number	Expiration																																																												
0.05	<u>SD3962</u>	<u>201707D</u>	<u>7/25/19</u>																																																												
0.08	<u>SD1013</u>	<u>201707E</u>	<u>7/25/19</u>																																																												
0.20	<u>SD3933</u>	<u>201707C</u>	<u>7/24/19</u>																																																												
0.08 DGS	N/A	<u>AG619605</u>	<u>7/14/18</u>																																																												
Simulator	Serial Number																																																														
0.00	<u>G2879</u>																																																														
Interferent	<u>G8144</u>																																																														
0.05	<u>G2078</u>																																																														
0.08	<u>G8149</u>																																																														
0.20	<u>G11621</u>																																																														
Attachments ☒ Form 41 ☒ Pre-Stability Tests ☐ Flow Calibration ☒ Calibration Adjustment ☒ Post-Stability Tests ☒ Other <u>form 40</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																																													
Notes/Suggested Service: <u>INSTRUMENT WILL NOT PRINT TO EXTERNAL</u> <u>PRINTER SUGGEST SENDING TO REPAIR</u> <u>QA/QC OK SP</u> <u>3/1/17</u> Quality Control Review		<u>9/14/17</u> Date																																																													

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
07/28/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:03
Control Test	0.049	08:03
Air Blank	0.000	08:04
Control Test	0.050	08:05
Air Blank	0.000	08:05
Control Test	0.050	08:06
Air Blank	0.000	08:06
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

P Murphy
Operator's Signature

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
07/28/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:58
Control Test	0.081	07:58
Air Blank	0.000	07:59
Control Test	0.081	08:00
Air Blank	0.000	08:00
Control Test	0.081	08:01
Air Blank	0.000	08:01
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
Operator's Signature

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
07/28/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:08
Control Test	0.192	08:09
Air Blank	0.000	08:10
Control Test	0.187	08:10
Air Blank	0.000	08:11
Control Test	0.187	08:12
Air Blank	0.000	08:12
Control Test Stats		
Average	0.1887	
Std Dev	0.0029	
Rel Std Dev(%)	1.5301	

P Murphy
Operator's Signature

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
07/28/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:17
Control Test	0.080	08:17
Air Blank	0.000	08:17
Control Test	0.080	08:18
Air Blank	0.000	08:18
Control Test	0.080	08:19
Air Blank	0.000	08:19
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

P Murphy
Operator's Signature

SP

9/14/17
JD

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
09/01/2017 11:28:02

Auto Calibration
Max Power Res Value = 41
Auto Range Res Value = 23

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 13047, Sum Io = 14222

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.1080 (-0.0120)
Sample #2 = 0.0740 (0.0330)
Sample #3 = 0.0970 (0.0400)
Sample #4 = 0.0920 (0.0810)
Avg % Abs = 0.0877 (0.0513)
STD DEV = 0.0121 (0.0259)
REL STD DEV = 13.799 (50.512)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.0830 (-0.0200)
Sample #2 = 0.0720 (-0.0070)
Sample #3 = 0.0890 (-0.0050)
Sample #4 = 0.1020 (-0.0110)
Avg % Abs = 0.0877 (-0.0077)
STD DEV = 0.0150 (0.0031)
REL STD DEV = 17.161 (39.848)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 13030, Sum Io = 14215

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.7720 (-0.0110)
Sample #2 = 0.7890 (-0.0120)
Sample #3 = 0.7820 (0.0100)
Sample #4 = 0.7760 (0.0370)
Avg % Abs = 0.7823 (0.0117)
STD DEV = 0.0065 (0.0245)
REL STD DEV = 0.832 (210.364)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.4490 (-0.0140)
Sample #2 = 1.4800 (-0.0270)
Sample #3 = 1.4700 (-0.0140)
Sample #4 = 1.4950 (-0.0250)
Avg % Abs = 1.4817 (-0.0220)
STD DEV = 0.0126 (0.0070)
REL STD DEV = 0.849 (31.818)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 13022, Sum Io = 14214

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.9000 (-0.0340)
Sample #2 = 1.8890 (-0.0080)
Sample #3 = 1.8710 (-0.0130)
Sample #4 = 1.8710 (-0.0050)
Avg % Abs = 1.8770 (-0.0087)
STD DEV = 0.0104 (0.0040)
REL STD DEV = 0.554 (46.632)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.5510 (-0.0130)
Sample #2 = 3.5570 (-0.0120)
Sample #3 = 3.5570 (-0.0120)
Sample #4 = 3.5580 (-0.0320)
Avg % Abs = 3.5573 (-0.0187)
STD DEV = 0.0006 (0.0115)
REL STD DEV = 0.016 (61.859)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 13019, Sum Io = 14213

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.5950 (-0.0240)
Sample #2 = 3.5970 (-0.0180)
Sample #3 = 3.5840 (0.0030)
Sample #4 = 3.6080 (-0.0060)
Avg % Abs = 3.5963 (-0.0070)
STD DEV = 0.0120 (0.0105)
REL STD DEV = 0.334 (150.509)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 13019, Sum Io = 14215

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 5.2900 (-0.0240)
Sample #2 = 5.2590 (0.0250)
Sample #3 = 5.2560 (0.0350)
Sample #4 = 5.2610 (0.0150)
Avg % Abs = 5.2587 (0.0250)
STD DEV = 0.0025 (0.0100)
REL STD DEV = 0.048 (40.000)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 9.7930 (0.0130)
Sample #2 = 9.7930 (0.0360)
Sample #3 = 9.7750 (0.0380)
Sample #4 = 9.7790 (0.0360)
Avg % Abs = 9.7823 (0.0367)
STD DEV = 0.0095 (0.0012)
REL STD DEV = 0.097 (3.149)

***** AUTO CAL DATA *****
Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.088
Std Dev = 0.01 Rel Std Dev = 13.80
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.782
Std Dev = 0.01 Rel Std Dev = 0.83
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.877
Std Dev = 0.01 Rel Std Dev = 0.55
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.596
Std Dev = 0.01 Rel Std Dev = 0.33
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.259
Std Dev = 0.00 Rel Std Dev = 0.05
Zero Order Coef = -198.95
First Order Coef = 2599.57
Second Order Coef = 29.30
Standard Deviation = 31.925846

SP

9/14/17
JD

<<<< CHANNEL 2 >>>>
 Sol Val = 0.000 mg/l or 0.000 g/210L
 % Abs = 0.088
 Std Dev = 0.02 Rel Std Dev = 17.16
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.482
 Std Dev = 0.01 Rel Std Dev = 0.85
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.557
 Std Dev = 0.00 Rel Std Dev = 0.02
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.752
 Std Dev = 0.01 Rel Std Dev = 0.10
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.782
 Std Dev = 0.01 Rel Std Dev = 0.10
 Zero Order Coef = -107.83
 First Order Coef = 1320.35
 Second Order Coef = 15.47
 Standard Deviation = 17.601114

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0006
0.040	0.039	0.0011
0.100	0.100	-0.0005
0.200	0.200	-0.0001
0.300	0.300	0.0001

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	0.0005
0.100	0.100	-0.0005
0.200	0.200	0.0002
0.300	0.300	-0.0001

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 3142.00
 Sample #2 = 3158.00
 Sample #3 = 3131.00
 Sample #4 = 3217.00
 Average Result = 3168.6667
 STD DEV = 43.9811
 REL STD DEV = 1.388

 ***** CHANNEL 2
 Sample #1 = 3353.00
 Sample #2 = 3358.00
 Sample #3 = 3368.00
 Sample #4 = 3399.00
 Average Result = 3375.0000
 STD DEV = 21.3776
 REL STD DEV = 0.633

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1012
 3 um H2O Adjust (mg/l*10,000) = 641
 9 um H2O Adjust (mg/l*10,000) = 434
 **** AUTO CAL PASS

80-001646

8P

9/14/17
 JD

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
09/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:31
Control Test	0.049	12:32
Air Blank	0.000	12:32
Control Test	0.049	12:33
Air Blank	0.000	12:33
Control Test	0.049	12:34
Air Blank	0.000	12:34
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P. Murphy
Operator's Signature

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
09/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:36
Control Test	0.080	12:37
Air Blank	0.000	12:37
Control Test	0.080	12:38
Air Blank	0.000	12:38
Control Test	0.080	12:39
Air Blank	0.000	12:39
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P. Murphy
Operator's Signature

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
09/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:41
Control Test	0.196	12:42
Air Blank	0.000	12:42
Control Test	0.198	12:43
Air Blank	0.000	12:43
Control Test	0.198	12:44
Air Blank	0.000	12:45
Control Test Stats		
Average	0.1973	
Std Dev	0.0012	
Rel Std Dev(%)	0.5852	

P. Murphy
Operator's Signature

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
09/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:46
Control Test	0.080	12:46
Air Blank	0.000	12:47
Control Test	0.080	12:47
Air Blank	0.000	12:48
Control Test	0.080	12:48
Air Blank	0.000	12:48
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

P. Murphy
Operator's Signature

Post Calibration - Adjust Stabilities

SP

9/14/17
JB