



INSTRUMENT PROCESSING SHEET

Agency Putnam County SO

S/N 80-001648

Florida Department of Law Enforcement

Date In 10/30/2017

DI Completion Date 11/2/17

Ship P/U H/D CMI EE

Intake Performed By TP

Annual
 Registration
 Return from CMI / EE

Visual Inspection:
 Case Handle
 Keyboard Dry Gas Shelf
 Feet Breath Tube
 Ports Screws Tight

Other Equipment/ Accessories:
 Power cord Printer Cable
 Static Bag 12V DC Cable

Notes: _____

Quality Checks Performed By TP

Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 162
 Flow Verification (L/s)
 Flow Column # ATP105
 32 mm 141 (.139 - .169)
 36 mm 152 (.156 - .190)
 53 mm 230 (.228 - .278)
 103 mm 507-499 (.447 - .547)
 Barometric Pressure Check
 Gauge ID # 28427
 Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	G2835	201707D 07/25/2019
0.080	DR1279	201707E 07/25/2019
0.200	SD1025	201707C 07/24/2019
0.080 DGS	N/A	AG626605 9/22/18

Flow Calibration Performed By TP

Flow Column # ATP105
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value 162
 Post Calibration Verification (L/s)
 Flow Column # ATP105
 32 mm 152 (.139 - .169)
 36 mm 167 (.156 - .190)
 53 mm 236 (.228 - .278)
 103 mm 506 (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By TP

Lab Temp °C 22.0
 External Digital Therm. ID#: 300501
 34°C +- .2 Serial #: G2835
 34°C +- .2 Serial #: DR1279
 34°C +- .2 Serial #: SD1025

Final Release Date

NOV 02 2017

Calibration Adjustment Performed By _____

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By TP

Barometric Pressure ID# 28427
 Gauge 1019 Instrument 1018
 Mouth Alcohol Solution Lot # 2016-c
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	G2880
Interferent	G2840
0.050	G2835
0.080	DR1279
0.200	SD1025

Attachments

Form 41 Flow Calibration
 Stability Checks Form 40
 Calibration Adjustment Other _____
 Post-Stability Checks

Notes/Suggested Service: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

TP 11/2/17 J. Decker 11/2/17
 Tech Review / Date Admin Review / Date

80-0016018
 Stability checks
 11/2/17

INTOXILYZER 8000
 Instrument Initialization
 05:33 11/02/2017

PUTNAM COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001648
 11/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:19
Control Test	0.049	06:19
Air Blank	0.000	06:20
Control Test	0.049	06:21
Air Blank	0.000	06:21
Control Test	0.049	06:22
Air Blank	0.000	06:22
Control Test	0.000	06:22
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

PUTNAM COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001648
 11/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:30
Control Test	0.078	06:30
Air Blank	0.000	06:31
Control Test	0.079	06:31
Air Blank	0.000	06:32
Control Test	0.080	06:32
Air Blank	0.000	06:33
Control Test	0.000	06:33
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

DGS

Operator's Signature

PUTNAM COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001648
 11/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:35
Control Test	0.200	06:35
Air Blank	0.000	06:36
Control Test	0.200	06:36
Air Blank	0.000	06:37
Control Test	0.200	06:37
Air Blank	0.000	06:38
Control Test	0.000	06:38
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Pjam

Operator's Signature

Operator's Signature

11/2/17
 [Signature]

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001648
11/02/2017
Software: 8100.27

Flow Rate Calibration*****

- 1: Rate (Liters/min) = 5
SQRT(Diff)) = 7.000
- 2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.957
- 3: Rate (Liters/min) = 30
SQRT(Diff)) = 22.359

Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 625
Rounded Intercept = -535693
Correlation = 0.99622

80-001648
Flow Cal.
11/2/17
@

fgm

11/2/17
22