

## INSTRUMENT PROCESSING SHEET

Agency NAS Jacksonville S/N 80-002097  
 Date In 8/25/17 Date Out 8/31/17  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Stable Bag</u> Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.5</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>212</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>A7P102</u> 32mm <u>148</u> (.139 - .169) 36mm <u>167</u> (.156 - .190) 53mm <u>234</u> (.228 - .278) 103mm <u>494</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3962</td> <td>201707P 7/25/19</td> </tr> <tr> <td>0.08</td> <td>SD1013</td> <td>201707E 7/25/19</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> <td>201707C 7/24/19</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605 7/14/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD3962	201707P 7/25/19	0.08	SD1013	201707E 7/25/19	0.20	SD3933	201707C 7/24/19	0.08 DGS	N/A	AG619605 7/14/18	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD3962	201707P 7/25/19															
0.08	SD1013	201707E 7/25/19															
0.20	SD3933	201707C 7/24/19															
0.08 DGS	N/A	AG619605 7/14/18															
		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Quality Checks Cont.</b> Performed By <u>[Signature]</u> <b>Simulator Temperatures °C</b> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1013</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3933</u>															

RECEIVED  
 AUG 31 2017  
 FDLE  
 Alcohol Testing Program

<b>Calibration Adjustment</b> Performed By _____ <input checked="" type="checkbox"/> Calibration Adjustment N/A <input type="checkbox"/> Calibration Adjustment Complete Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			<b>Department Inspection</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1014</u> Gauge ID# <u>26932</u> <u>1013</u> Instrument Mouth Alcohol Solution Lot # <u>206-c</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD3962</td> </tr> <tr> <td>0.08</td> <td>SD1013</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD3962	0.08	SD1013	0.20	SD3933
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1019																																																												
Interferent	SD1021																																																												
0.05	SD3962																																																												
0.08	SD1013																																																												
0.20	SD3933																																																												

<b>Attachments</b>	
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____

Notes/Suggested Service:  
QA/QC OK 8/31/2017  
 \_\_\_\_\_  
 \_\_\_\_\_  
[Signature] 8/31/17  
 Quality Control Review Date

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
---

80-002097  
 Stability Checks  
 8/31/17

INTOXILYZER 8000  
 Instrument Initialization  
 07:51 08/31/2017

NAS JACKSONVILLE PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-002097  
 08/31/2017  
 Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	08:44
Control Test	0.080	08:44
Air Blank	0.000	08:45
Control Test	0.080	08:45
Air Blank	0.000	08:46
Control Test	0.079	08:46
Air Blank	0.000	08:47
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DSS

Operator's Signature

8/31/17  
 JS

NAS JACKSONVILLE PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-002097  
 08/31/2017  
 Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	08:51
Control Test	0.081	08:52
Air Blank	0.000	08:53
Control Test	0.080	08:53
Air Blank	0.000	08:54
Control Test	0.081	08:55
Air Blank	0.000	08:55
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Operator's Signature

NAS JACKSONVILLE PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-002097  
 08/31/2017  
 Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	08:57
Control Test	0.050	08:57
Air Blank	0.000	08:58
Control Test	0.049	08:59
Air Blank	0.000	08:59
Control Test	0.050	09:00
Air Blank	0.000	09:00
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Operator's Signature

NAS JACKSONVILLE PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-002097  
 08/31/2017  
 Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	09:02
Control Test	0.202	09:03
Air Blank	0.000	09:03
Control Test	0.200	09:04
Air Blank	0.000	09:05
Control Test	0.199	09:05
Air Blank	0.000	09:06
Control Test Stats		
Average	0.2003	
Std Dev	0.0015	
Rel Std Dev(%)	0.7625	

ASB

Operator's Signature