



# INSTRUMENT PROCESSING SHEET

Agency New Smyrna Beach PDS/N 80-002231Florida Department of  
Law EnforcementDate In 10/24/17DI Completion Date 10/27/17 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ <b>Final Release Date</b>  <div style="text-align: center; font-size: 24pt; font-weight: bold;">OCT 31 2017</div>	<b>Quality Checks</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>229</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.158</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>62835</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AE626605</u> <u>9/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AE626605</u> <u>9/22/18</u>	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.2</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>62835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>
Simulator	Serial #	Lot #/Exp															
0.050	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>															
0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>															
0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>															
0.080 DGS	N/A	<u>AE626605</u> <u>9/22/18</u>															

<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>[Signature]</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1016</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>62830</u></td> </tr> <tr> <td>Interferent</td> <td><u>62840</u></td> </tr> <tr> <td>0.050</td> <td><u>62835</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>62830</u>	Interferent	<u>62840</u>	0.050	<u>62835</u>	0.080	<u>DR1279</u>	0.200	<u>SD1025</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	<u>62830</u>																																																												
Interferent	<u>62840</u>																																																												
0.050	<u>62835</u>																																																												
0.080	<u>DR1279</u>																																																												
0.200	<u>SD1025</u>																																																												

Notes/Suggested Service: _____ _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  <u>[Signature]</u> 10/30/17 <u>[Signature]</u> 10/31/17 Tech Review / Date      Admin Review / Date
---	---

NEW SMYRNA BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002231  
10/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:00
Control Test	0.050	11:00
Air Blank	0.000	11:01
Control Test	0.050	11:02
Air Blank	0.000	11:02
Control Test	0.050	11:03
Air Blank	0.000	11:03
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*P Murphy*  
-----  
Operator's Signature

NEW SMYRNA BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002231  
10/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:53
Control Test	0.081	10:54
Air Blank	0.000	10:54
Control Test	0.081	10:55
Air Blank	0.000	10:55
Control Test	0.081	10:56
Air Blank	0.000	10:57
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*P Murphy*  
-----  
Operator's Signature

NEW SMYRNA BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002231  
10/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:50
Control Test	0.199	09:50
Air Blank	0.000	09:51
Control Test	0.200	09:52
Air Blank	0.000	09:52
Control Test	0.201	09:53
Air Blank	0.000	09:53
Control Test Stats		
Average	0.2000	
Std Dev	0.0010	
Rel Std Dev(%)	0.5000	

*P Murphy*  
-----  
Operator's Signature

NEW SMYRNA BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002231  
10/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:05
Control Test	0.079	11:05
Air Blank	0.000	11:06
Control Test	0.079	11:06
Air Blank	0.000	11:07
Control Test	0.079	11:07
Air Blank	0.000	11:07
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

*P Murphy*  
-----  
Operator's Signature

SP

10/31/17  
JD