



Alcohol Testing Program

### INSTRUMENT PROCESSING SHEET

Agency Walla Walla PD

S/N 80-002989

Date In 3/29/17

Date Out 4/6/17

Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:

Case  Handle  
 Dry Gas Holder  Feet  
 Keyboard/Plug  Back/Plugs  
 Screws tight  Breath Hose

Other Equipment:

Power cord  
 Printer Cable  
 Other \_\_\_\_\_

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 208  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm .156 (.139 - .169)  
 36mm .160 (.156 - .190)  
 53mm .230 (.228 - .278)  
 103mm .503 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	DR2035	2016030 3/8/18
0.08	SD3964	2016118 11/15/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG626604 9/22/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

\_\_\_\_\_

\_\_\_\_\_

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete  
 Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1011 Gauge  
 ID# 28421 1009 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	DR2035
0.08	SD3964
0.20	SD1025

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: OK [Signature] 4/6/17

\_\_\_\_\_

\_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kvickland

Quality Control Review

4/6/17  
Date

TALLAHASSEE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002989  
04/06/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:55
Control Test	0.049	08:56
Air Blank	0.000	08:56
Control Test	0.049	08:57
Air Blank	0.000	08:58
Control Test	0.050	08:58
Air Blank	0.000	08:59
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

*P Murphy*  
Operator's Signature

TALLAHASSEE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002989  
04/06/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:01
Control Test	0.080	09:02
Air Blank	0.000	09:02
Control Test	0.080	09:03
Air Blank	0.000	09:03
Control Test	0.079	09:04
Air Blank	0.000	09:05
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

*P Murphy*  
Operator's Signature

TALLAHASSEE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002989  
04/06/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:06
Control Test	0.195	09:07
Air Blank	0.000	09:08
Control Test	0.196	09:08
Air Blank	0.000	09:09
Control Test	0.197	09:10
Air Blank	0.000	09:10
Control Test Stats		
Average	0.1960	
Std Dev	0.0010	
Rel Std Dev(%)	0.5102	

*P Murphy*  
Operator's Signature

TALLAHASSEE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002989  
04/06/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:14
Control Test	0.079	09:14
Air Blank	0.000	09:15
Control Test	0.080	09:15
Air Blank	0.000	09:15
Control Test	0.079	09:16
Air Blank	0.000	09:16
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

*P Murphy*  
Operator's Signature

*[Signature]* JSK