



Alcohol Testing Program

### INSTRUMENT PROCESSING SHEET

Agency FWC - Ocala

S/N 80-003195

Date In 3/27/17

Date Out 4/11/17

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input checked="" type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>120</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>A7P105</u> 32mm <u>142</u> (.139 - .169) 36mm <u>160</u> (.156 - .190) 53mm <u>233</u> (.228 - .278) 103mm <u>479</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>DR2035</u></td> <td><u>201603D</u> <u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> <td><u>201611B</u> <u>11/15/18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> <td><u>201604C</u> <u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG626604</u> <u>7/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>	0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>	0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>	0.08 DGS	N/A	<u>AG626604</u> <u>7/22/18</u>	<b>Flow Calibration</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>															
0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>															
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>															
0.08 DGS	N/A	<u>AG626604</u> <u>7/22/18</u>															
		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____															

RECEIVED  
APR 11 2017  
FDLE  
Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____			
<input checked="" type="checkbox"/> Optical Bench Calibration N/A			
<input type="checkbox"/> Optical Bench Calibration Complete			
Barometric Pressure Gauge _____ ID # _____			
Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		
<input type="checkbox"/> Post Calibration Stability Checks			
Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

<b>Department Inspection</b> Performed By <u>[Signature]</u>	
<input type="checkbox"/> Barometric Pressure ID# <u>28421</u>	<u>1021</u> Gauge <u>1022</u> Instrument
Mouth Alcohol Solution Lot # <u>2016-C</u>	
Acetone Stock Solution Lot # <u>2017-A</u>	
Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD3964</u>
0.20	<u>SD1025</u>
<b>Attachments</b>	
<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: QC RMB  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC
<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC
<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use
<input type="checkbox"/> Remain Out of Evidentiary Use
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review

4/11/17  
Date

80-003195  
 Stability checks  
 4/11/17

INTOXILYZER 8000  
 Instrument Initialization  
 08:22 04/11/2017

FWC  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003195  
 04/11/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:57
Control Test	0.079	08:58
Air Blank	0.000	08:58
Control Test	0.079	08:59
Air Blank	0.000	09:00
Control Test	0.079	09:00
Air Blank	0.000	09:01
Control Test	0.000	09:01
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FWC  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003195  
 04/11/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:03
Control Test	0.049	09:04
Air Blank	0.000	09:04
Control Test	0.049	09:05
Air Blank	0.000	09:05
Control Test	0.049	09:06
Air Blank	0.000	09:07
Control Test	0.000	09:07
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FWC  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003195  
 04/11/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:08
Control Test	0.080	09:08
Air Blank	0.000	09:09
Control Test	0.080	09:09
Air Blank	0.000	09:10
Control Test	0.081	09:10
Air Blank	0.000	09:10
Control Test	0.000	09:10
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS  
 B15

Operator's Signature

ADS

Operator's Signature