



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FWC - Ocala

S/N 80-003196

Date In 3/27/17

Date Out 5/16/17

Ship P/U H/D CMI EE

Intake Performed By TA **Quality Checks** Performed By TA **Flow Calibration** Performed By TA

- Registration
- Annual
- Return from CMI
- Return from Enforcement Electronics
- Other _____

Visual Inspection:

Case OK Handle OK

OK Dry Gas Holder OK Feet OK

OK Keyboard/Plug OK Back/Plugs OK

OK Screws tight OK Breath Hose OK

- Other Equipment:
- Power cord
 - Printer Cable
 - Other _____

Notes: Exhaust cover broken will not attach

- Lab Temp °C 21.0
- Breath Tube Screen
- Replace External O-Rings
- Instrument Set Up Verified
- R-Value 139
- Flow Verification (L/s)

Flow Column # ATP103

32mm 152 (.139 - .169)

36mm 161 (.156 - .190)

53mm 233 (.228 - .278)

103mm 501 (.447 - .547)

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|---------------------|
| 0.05 | G3709 | 201603D 3/8/18 |
| 0.08 | DR1279 | 201611B 11/15/18 |
| 0.20 | DR3856 | 201604C 4/5/18 |
| 0.08 DGS | N/A | AS626604 9/22/18 |

- Flow Calibration N/A
- Flow Calibration Complete
- Flow Column # _____
- 5L/min - 50mm
- 15L/min - 53mm
- 30L/min - 103mm
- R-Value _____
- Post Calibration Verification (L/s)

- Maintenance** Performed By _____
- Battery Replacement
 - Dry Gas Regulator Replacement
 - Breath Tube Replacement
 - Other _____

Quality Checks Cont. Performed By TA

Simulator Temperatures °C

External Digital Therm. ID #: 300505

- 34°C +/- .2 Serial #: G3709
- 34°C +/- .2 Serial #: DR1279
- 34°C +/- .2 Serial #: DR3856

Calibration Adjustment Performed By _____

- Calibration Adjustment N/A
- Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.300 | | | |
| 0.080 DGS | N/A | | |

Post Calibration Adjustment Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Department Inspection Performed By TA

- Barometric Pressure 1015 Gauge
- ID# 26932 1014 Instrument

Mouth Alcohol Solution Lot # 2016-C

Acetone Stock Solution Lot # 2017-A

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | G2880 |
| Interferent | G2834 |
| 0.05 | G3709 |
| 0.08 | DR1279 |
| 0.20 | DR3856 |

Attachments

- Form 41
- Pre-Stability Tests
- Flow Calibration
- Calibration Adjustment
- Post-Stability Tests
- Other Form 40

Notes/Suggested Service:

Please change level 2 P.W. to something unique.

Exhaust Port Cover changed.

TA/OC: [Signature]

Instrument Complies with Chapter 11D-8, FAC

Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use

Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date 5/16/17

RECEIVED
MAY 16 2017
FDLE
Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FWC - Ocala S/N 80-003196
 Date In 3/27/17 Date Out _____
 Ship P/U H/D CMI EE

Intake Performed By JD

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case OK Handle OK
OK Dry Gas Holder OK Feet OK
OK Keyboard/Plug OK Back/Plugs OK
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: Exhaust Cover Broken. Will not attach
JD

Quality Checks Performed By _____

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value _____
 Flow Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Barometric Pressure Check
 Gauge ID # _____

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|-----------|
| 0.05 | | |
| 0.08 | | |
| 0.20 | | |
| 0.08 DGS | N/A | |

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 47mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.400 | | | |
| 0.080 DGS | N/A | | |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Department Inspection Performed By _____

Barometric Pressure ID# _____ Gauge _____ Instrument _____

Mouth Alcohol Solution Lot # _____
 Acetone Stock Solution Lot # _____

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | |
| Interferent | |
| 0.05 | |
| 0.08 | |
| 0.20 | |

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: _____
QC: RDB
JD 5/16/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

RECEIVED
 MAY 16 2017
 FDLE
 Alcohol Testing Program

80-003196
 Stability Checks
 5/16/17
 5/16/17

.INTOXILYZER 8000
 Instrument Initialization
 07:31 05/16/2017

FIC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003196
 05/16/2017
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:27 |
| Control Test | 0.078 | 08:27 |
| Air Blank | 0.000 | 08:28 |
| Control Test | 0.078 | 08:28 |
| Air Blank | 0.000 | 08:29 |
| Control Test | 0.078 | 08:29 |
| Air Blank | 0.000 | 08:29 |
| Control Test Stats | | |
| Average | 0.0780 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

DAS

Operator's Signature

JD

FIC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003196
 05/16/2017
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:31 |
| Control Test | 0.048 | 08:32 |
| Air Blank | 0.000 | 08:32 |
| Control Test | 0.048 | 08:33 |
| Air Blank | 0.000 | 08:34 |
| Control Test | 0.048 | 08:34 |
| Air Blank | 0.000 | 08:35 |
| Control Test Stats | | |
| Average | 0.0480 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

Operator's Signature

FIC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003196
 05/16/2017
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:36 |
| Control Test | 0.194 | 08:37 |
| Air Blank | 0.000 | 08:38 |
| Control Test | 0.194 | 08:38 |
| Air Blank | 0.000 | 08:39 |
| Control Test | 0.194 | 08:40 |
| Air Blank | 0.000 | 08:40 |
| Control Test Stats | | |
| Average | 0.1940 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

Operator's Signature

FIC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003196
 05/16/2017
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:41 |
| Control Test | 0.078 | 08:42 |
| Air Blank | 0.000 | 08:43 |
| Control Test | 0.078 | 08:43 |
| Air Blank | 0.000 | 08:44 |
| Control Test | 0.078 | 08:45 |
| Air Blank | 0.000 | 08:45 |
| Control Test Stats | | |
| Average | 0.0780 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

Operator's Signature

DAS