

INSTRUMENT PROCESSING SHEET

Agency FHP - Ocala S/N 80-003407
 Date In 1/13/17 Date Out 2/2/17 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 203
 Flow Verification (L/s)
 Flow Column # A7P102
 32mm 158 (.139 - .169)
 36mm 160 (.156 - .190)
 53mm 261 (.228 - .278)
 103mm 517 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	201603D 3/8/18
0.08	G8149	201601F 1/26/18
0.20	G11621	201604C 4/15/18
0.08 DGS	N/A	A6626605 9/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 FEB 02 2017
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure Gauge ID# 28421 1020 1021

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G11739
0.08	G8149
0.20	G11621

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: [Signature]

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

[Signature]
 Quality Control Review

2/2/17
 Date

80-003407
 Stability Checks
 2/2/17

INTOXILYZER 8000
 Instrument Initialization
 08:31 02/02/2017

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003407
 02/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:05
Control Test	0.201	09:06
Air Blank	0.000	09:06
Control Test	0.200	09:07
Air Blank	0.000	09:07
Control Test	0.200	09:08
Air Blank	0.000	09:09
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003407
 02/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:09
Control Test	0.080	09:10
Air Blank	0.000	09:11
Control Test	0.080	09:11
Air Blank	0.000	09:12
Control Test	0.080	09:13
Air Blank	0.000	09:13
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003407
 02/02/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:14
Control Test	0.050	09:15
Air Blank	0.000	09:15
Control Test	0.050	09:16
Air Blank	0.000	09:16
Control Test	0.050	09:17
Air Blank	0.000	09:18
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

DGS

Operator's Signature

TSK
 RDB