

Agency FL Highway Patrol S/N 80-003411
 Date In 3/24/17 Date Out 3/24/17

MAR 27 2017

Ship P/U H/D CMI FDLE

Alcohol Testing Program

Intake Performed By <u>ONS</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: OK Case OK Handle OK Dry Gas Holder OK Feet OK Keyboard/Plug OK Back/Plugs OK Screws tight OK Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>ONS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>199</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>0.160</u> (.139 - .169) 36mm <u>0.175</u> (.156 - .190) 53mm <u>0.246</u> (.228 - .278) 103mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>G3709</u></td> <td><u>201603D</u> <u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>DR1279</u></td> <td><u>201601F</u> <u>1/26/18</u></td> </tr> <tr> <td>0.20</td> <td><u>DR3856</u></td> <td><u>201604C</u> <u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG1626605</u> <u>9/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>G3709</u>	<u>201603D</u> <u>3/8/18</u>	0.08	<u>DR1279</u>	<u>201601F</u> <u>1/26/18</u>	0.20	<u>DR3856</u>	<u>201604C</u> <u>4/5/18</u>	0.08 DGS	N/A	<u>AG1626605</u> <u>9/22/18</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>G3709</u>	<u>201603D</u> <u>3/8/18</u>															
0.08	<u>DR1279</u>	<u>201601F</u> <u>1/26/18</u>															
0.20	<u>DR3856</u>	<u>201604C</u> <u>4/5/18</u>															
0.08 DGS	N/A	<u>AG1626605</u> <u>9/22/18</u>															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.400																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

Department Inspection Performed By <u>ONS</u> <input checked="" type="checkbox"/> Barometric Pressure _____ Gauge ID# <u>28427</u> _____ Instrument	
Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2016-B</u>	
Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>G3709</u>
0.08	<u>DR1279</u>
0.20	<u>DR3856</u>
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	

Notes: _____

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

OC Q 3/27/17
 Quality Control Review

Scott Kirkland 3/27/17
 Date

Stability Checks # 80-003411 FL Highway Patrol 3/24/17 *RMS*

RMS

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003411
03/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:52
Control Test	0.050	13:53
Air Blank	0.000	13:54
Control Test	0.050	13:54
Air Blank	0.000	13:55
Control Test	0.050	13:55
Air Blank	0.000	13:56
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003411
03/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:57
Control Test	0.078	13:58
Air Blank	0.000	13:59
Control Test	0.079	13:59
Air Blank	0.000	14:00
Control Test	0.079	14:01
Air Blank	0.000	14:01
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003411
03/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:02
Control Test	0.196	14:03
Air Blank	0.000	14:04
Control Test	0.196	14:04
Air Blank	0.000	14:05
Control Test	0.196	14:06
Air Blank	0.000	14:06
Control Test Stats		
Average	0.1960	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003411
03/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:08
Control Test	0.081	14:08
Air Blank	0.000	14:09
Control Test	0.081	14:09
Air Blank	0.000	14:09
Control Test	0.080	14:10
Air Blank	0.000	14:10
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

SK

SK

RMS
Operator's Signature

RMS
Operator's Signature

RMS
Operator's Signature

RMS
Operator's Signature