



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Osceola County SO

S/N 80-003935

Date In 3/2/17

Date Out 3/15/17

Ship P/U H/D CMI EE

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By TP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 123
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm 0.144 (.139 - .169)
 36mm 0.164 (.156 - .190)
 53mm 0.238 (.228 - .278)
 103mm 0.511 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	<u>N/A</u>	<u>AG626605</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
MAR 16 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

Department Inspection Performed By TP

Barometric Pressure 1025 Gauge
 ID# 26932 1025 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>SD39 DR2035</u>
0.08	<u>SD3964</u>
0.20	<u>SD1025</u>

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: Please change level 2 menu password to something unique. @DS
QA/QC OK SP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kuitland

Quality Control Review

3/16/17
Date

Stability Checks # 80-003935 Osceola County S.O. 3/15/17 ~~RMS~~

~~RMS~~

OSCEOLA COUNTY S.O.
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-003935
03/15/2017
Software: 8100.27

Test	9/21/01	Time
Air Blank	0.000	13:04
Control Test	0.050	13:04
Air Blank	0.000	13:05
Control Test	0.050	13:06
Air Blank	0.000	13:06
Control Test	0.050	13:07
Air Blank	0.000	13:07
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

~~RMS~~

Operator's Signature

OSCEOLA COUNTY S.O.
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-003935
03/15/2017
Software: 8100.27

Test	9/21/01	Time
Air Blank	0.000	12:50
Control Test	0.198	12:51
Air Blank	0.000	12:51
Control Test	0.090	12:52
Air Blank	0.000	12:53
Control Test	0.091	12:53
Air Blank	0.000	12:54
Control Test Stats		
Average	0.0907	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

~~RMS~~

Operator's Signature

OSCEOLA COUNTY S.O.
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-003935
03/15/2017
Software: 8100.27

Test	9/21/01	Time
Air Blank	0.000	12:55
Control Test	0.198	12:56
Air Blank	0.000	12:56
Control Test	0.198	12:57
Air Blank	0.000	12:57
Control Test	0.198	12:58
Air Blank	0.000	12:59
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

~~RMS~~

Operator's Signature

OSCEOLA COUNTY S.O.
Intoxilyzer - Alconol Analyzer
Model 8000 SN 80-003935
03/15/2017
Software: 8100.27

Test	9/21/01	Time
Air Blank	0.000	13:00
Control Test	0.075	13:00
Air Blank	0.000	13:01
Control Test	0.080	13:01
Air Blank	0.000	13:02
Control Test	0.080	13:02
Air Blank	0.000	13:03
Control Test Stats		
Average	0.0757	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

RK SP

~~RMS~~

Operator's Signature