



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Sarasota County S/N 80-005076

Date In 3/10/17 Date Out 3/30/17 Ship P/U H/D CMI EE

Intake Performed By DP

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 222
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .156 (.139 - .169)
 36mm .171 (.156 - .190)
 53mm .234 (.228 - .278)
 103mm .511 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G3709</u>	<u>201603D</u> <u>3-8-18</u>
0.08	<u>DR1279</u>	<u>20161FB</u> <u>11-15-18</u>
0.20	<u>DR3556</u>	<u>201604C</u> <u>4-5-18</u>
0.08 DGS	N/A	<u>AG1626604</u> <u>9-22-18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
APR 03 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure ID# 28427 Gauge 1013 Instrument 1011

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>G3709</u>
0.08	<u>DR1279</u>
0.20	<u>DR3556</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: Please change level 2 password to something unique. SP
QC: DP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review

4/3/17
Date

SARASOTA COUNTY SO - STABILITY CHECKS - INSTRUMENT # 80-005076

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005076
 03/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:48
Control Test	0.049	09:49
Air Blank	0.000	09:49
Control Test	0.048	09:50
Air Blank	0.000	09:50
Control Test	0.049	09:51
Air Blank	0.000	09:52
Control Test Stats	0.0487	
Average	0.0006	
Std Dev	1.1863	
Rel Std Dev(%)		

SP

Operator's Signature

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005076
 03/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:53
Control Test	0.078	09:54
Air Blank	0.000	09:54
Control Test	0.078	09:55
Air Blank	0.000	09:55
Control Test	0.078	09:56
Air Blank	0.000	09:57
Control Test Stats	0.0780	
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005076
 03/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:39
Control Test	0.198	09:40
Air Blank	0.000	09:40
Control Test	0.198	09:41
Air Blank	0.000	09:42
Control Test	0.198	09:42
Air Blank	0.000	09:43
Control Test Stats	0.1980	
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005076
 03/30/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:35
Control Test	0.080	09:35
Air Blank	0.000	09:35
Control Test	0.080	09:36
Air Blank	0.000	09:36
Control Test	0.080	09:36
Air Blank	0.000	09:37
Control Test Stats	0.0800	
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

AKNS BK

SP

Operator's Signature