



Alcohol Testing Program

## INSTRUMENT PROCESSING SHEET

Agency Hillsborough CountyS/N 80-005113Date In 5/9/13Date Out 5/17/17☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   |  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |  |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|
| <b>Intake</b> Performed By <u>QAC</u><br><input type="checkbox"/> Registration<br><input checked="" type="checkbox"/> Annual<br><input type="checkbox"/> Return from CMI<br><input type="checkbox"/> Return from Enforcement Electronics<br><input type="checkbox"/> Other _____<br>Visual Inspection:<br><u>OK</u> Case <u>OK</u> Handle<br><u>OK</u> Dry Gas Holder <u>OK</u> Feet<br><u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs<br><u>OK</u> Screws tight <u>OK</u> Breath Hose<br>Other Equipment:<br><input type="checkbox"/> Power cord<br><input type="checkbox"/> Printer Cable<br><input checked="" type="checkbox"/> Other <u>Static Bag</u><br>Notes: _____<br>_____<br>_____<br>_____ |  | <b>Quality Checks</b> Performed By <u>QAC</u><br><input checked="" type="checkbox"/> Lab Temp °C <u>21.8</u><br><input checked="" type="checkbox"/> Breath Tube Screen<br><input checked="" type="checkbox"/> Replace External O-Rings<br><input checked="" type="checkbox"/> Instrument Set Up Verified<br><input checked="" type="checkbox"/> R-Value <u>231</u><br><input checked="" type="checkbox"/> Flow Verification (L/s)<br>Flow Column # <u>ATP103</u><br>32mm <u>0.148</u> (.139 - .169)<br>36mm <u>0.164</u> (.156 - .190)<br>53mm <u>0.242</u> (.228 - .278)<br>103mm <u>0.533</u> (.447 - .547)<br><input checked="" type="checkbox"/> Barometric Pressure Check<br>Gauge ID # <u>26932</u> <u>INS 1004</u><br><input checked="" type="checkbox"/> Stability Checks |  | <b>Flow Calibration</b> Performed By <u>QAC</u><br><input checked="" type="checkbox"/> Flow Calibration N/A<br><input type="checkbox"/> Flow Calibration Complete<br>Flow Column # _____<br><input type="checkbox"/> 5L/min - 17mm<br><input type="checkbox"/> 15L/min - 53mm<br><input type="checkbox"/> 30L/min - 103mm<br><input type="checkbox"/> R-Value _____<br><input type="checkbox"/> Post Calibration Verification (L/s)<br>Flow Column # _____<br>32mm _____ (.139 - .169)<br>36mm _____ (.156 - .190)<br>53mm _____ (.228 - .278)<br>103mm _____ (.447 - .547) |  |
|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |  | <b>Maintenance</b> Performed By _____<br><input type="checkbox"/> Battery Replacement<br><input type="checkbox"/> Dry Gas Regulator Replacement<br><input type="checkbox"/> Breath Tube Replacement<br><input type="checkbox"/> Other _____                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |  |
|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |  | <b>Quality Checks Cont.</b> Performed By <u>QAC</u><br><b>Simulator Temperatures °C</b><br>External Digital Therm. ID #: <u>300505</u><br><input checked="" type="checkbox"/> 34°C +-.2 Serial #: <u>G3709</u><br><input checked="" type="checkbox"/> 34°C +-.2 Serial #: <u>DR1279</u><br><input checked="" type="checkbox"/> 34°C +-.2 Serial #: <u>DR3856</u>                                                                                                                                                                                                                                                                                                                                                                                                                  |  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |  |

RECEIVED  
MAY 23 2017  
FDLE  
Alcohol Testing Program

300505  
SD3962  
DR1279  
DR3856

## Calibration Adjustment

Performed By QAC

- ☐
- Calibration Adjustment N/A
- 
- ☒
- Calibration Adjustment Complete

Barometric Pressure Gauge 1016 ID # 28427

| Simulator | Serial Number | Lot Number        | Expiration      |
|-----------|---------------|-------------------|-----------------|
| 0.000     | <u>SD1016</u> | <u>N/A</u>        | <u>N/A</u>      |
| 0.040     | <u>SD1024</u> | <u>16320</u>      | <u>10/21/18</u> |
| 0.100     | <u>G11739</u> | <u>17060</u>      | <u>2/14/19</u>  |
| 0.200     | <u>SD1012</u> | <u>17090</u>      | <u>2/24/19</u>  |
| 0.300     | <u>G4444</u>  | <u>16410</u>      | <u>12/19/18</u> |
| 0.080 DGS | <u>N/A</u>    | <u>15615080A2</u> | <u>7/5/17</u>   |

## Post Calibration Adjustment Stability Checks

| Simulator          | Serial Number | Lot Number      | Expiration      |
|--------------------|---------------|-----------------|-----------------|
| 0.05 <u>SD3962</u> | <u>G3709</u>  | <u>201603D</u>  | <u>3/8/18</u>   |
| 0.08               | <u>DR1279</u> | <u>201611B</u>  | <u>11/15/18</u> |
| 0.20               | <u>DR3856</u> | <u>201604C</u>  | <u>4/5/18</u>   |
| 0.08 DGS           | <u>N/A</u>    | <u>AG626604</u> | <u>9/22/18</u>  |

## Department Inspection

Performed By QAC

- ☒
- Barometric Pressure
- 
- ID #
- 28421
- 1016
- Gauge
- 
- 1016
- Instrument

Mouth Alcohol Solution Lot # 2016-CAcetone Stock Solution Lot # 2017-A

| Simulator   | Serial Number |
|-------------|---------------|
| 0.00        | <u>G2880</u>  |
| Interferent | <u>G2834</u>  |
| 0.05        | <u>SD3962</u> |
| 0.08        | <u>DR1279</u> |
| 0.20        | <u>DR3856</u> |

## Attachments

- ☒
- Form 41
- 
- ☒
- Pre-Stability Tests
- 
- ☐
- Flow Calibration
- 
- ☒
- Calibration Adjustment
- 
- ☒
- Post-Stability Tests
- 
- ☐
- Other \_\_\_\_\_

## Notes/Suggested Service:

Performed optical bench calibration adjustment  
to bring barometric pressure value closer to nominal.  
QA/QC SP

JJ Graham

5/23/17

Quality Control Review

Date

- ☒
- Instrument Complies with Chapter 11D-8, FAC
- 
- ☐
- Instrument Does Not Comply with Chapter 11D-8, FAC
- 
- ☒
- Return to/Place into Evidentiary Use
- 
- ☐
- Remain Out of Evidentiary Use
- 
- ☒
- Conduct an Agency Inspection Before Evidentiary Use

# Pre-Calibration Adjustment

Stability Checks #80-005113 Hillsborough County SO. 5/16/17 ~~SP~~

*SP*

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
05/16/2017  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
05/16/2017  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
05/16/2017  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
05/16/2017  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 11:40 |
| Control Test       | 0.049  | 11:41 |
| Air Blank          | 0.000  | 11:41 |
| Control Test       | 0.049  | 11:42 |
| Air Blank          | 0.000  | 11:42 |
| Control Test       | 0.050  | 11:43 |
| Air Blank          | 0.000  | 11:44 |
| Control Test Stats |        |       |
| Average            | 0.0493 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 1.1703 |       |

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 11:45 |
| Control Test       | 0.077  | 11:46 |
| Air Blank          | 0.000  | 11:46 |
| Control Test       | 0.079  | 11:47 |
| Air Blank          | 0.000  | 11:48 |
| Control Test       | 0.079  | 11:48 |
| Air Blank          | 0.000  | 11:49 |
| Control Test Stats |        |       |
| Average            | 0.0783 |       |
| Std Dev            | 0.0012 |       |
| Rel Std Dev(%)     | 1.4741 |       |

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 11:51 |
| Control Test       | 0.195  | 11:51 |
| Air Blank          | 0.000  | 11:52 |
| Control Test       | 0.196  | 11:53 |
| Air Blank          | 0.000  | 11:53 |
| Control Test       | 0.197  | 11:54 |
| Air Blank          | 0.000  | 11:55 |
| Control Test Stats |        |       |
| Average            | 0.1960 |       |
| Std Dev            | 0.0010 |       |
| Rel Std Dev(%)     | 0.5102 |       |

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 11:56 |
| Control Test       | 0.079  | 11:57 |
| Air Blank          | 0.000  | 11:57 |
| Control Test       | 0.078  | 11:57 |
| Air Blank          | 0.000  | 11:58 |
| Control Test       | 0.078  | 11:58 |
| Air Blank          | 0.000  | 11:59 |
| Control Test Stats |        |       |
| Average            | 0.0783 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 0.7370 |       |

*SP*

Operator's Signature

*JD*

*SP*

Operator's Signature

*SP*

Operator's Signature

*SP*

Operator's Signature



# Optical Bench Calibration Adjustment Data #80-005113 Hillsborough County S.O. 5/17/17 RBS

HILLSBOROUGH CO SO  
Intoxilyzer - Alconol Analyzer  
Model 8000  
05/17/2017  
SN 80-005113  
10:19:47

## Auto Calibration

Max Power Res Value = 28  
Auto Range Res Value = 7

Sol Value = 0.000 g/210L \*\*\*

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 11955, Sum Io = 14237

Sum Io = 11932, Sum Io = 14228

Sample % Abs (% Abs Ref)

Sample #1 = 0.1350 (-0.0270)

Sample #2 = 0.0850 (0.0320)

Sample #3 = 0.0760 (0.0660)

Sample #4 = 0.0970 (0.0670)

Aug % Abs = 0.0860 (0.0550)

STD DEV = 0.0105 (0.0199)

REL STD DEV = 12.251 (36.227)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 0.1120 (-0.0140)

Sample #2 = 0.0770 (0.0140)

Sample #3 = 0.0770 (0.0280)

Sample #4 = 0.0910 (0.0140)

Aug % Abs = 0.0817 (0.0187)

STD DEV = 0.0081 (0.0081)

REL STD DEV = 9.897 (43.301)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 0.0680 (0.0020)

Sample #2 = 0.0960 (0.0130)

Sample #3 = 0.0490 (0.0550)

Sample #4 = 0.0430 (0.0780)

Aug % Abs = 0.0627 (0.0487)

STD DEV = 0.0290 (0.0330)

REL STD DEV = 3.364 (67.725)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 1.5650 (-0.0110)

Sample #2 = 1.5720 (0.0100)

Sample #3 = 1.5630 (0.0210)

Sample #4 = 1.5470 (0.0440)

Aug % Abs = 1.5607 (0.0250)

STD DEV = 0.0127 (0.0173)

REL STD DEV = 0.811 (69.397)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 1.5650 (-0.0110)

Sample #2 = 1.5720 (0.0100)

Sample #3 = 1.5630 (0.0210)

Sample #4 = 1.5470 (0.0440)

Aug % Abs = 1.5607 (0.0250)

STD DEV = 0.0127 (0.0173)

REL STD DEV = 0.811 (69.397)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 1.5650 (-0.0110)

Sample #2 = 1.5720 (0.0100)

Sample #3 = 1.5630 (0.0210)

Sample #4 = 1.5470 (0.0440)

Aug % Abs = 1.5607 (0.0250)

STD DEV = 0.0127 (0.0173)

REL STD DEV = 0.811 (69.397)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 1.5650 (-0.0110)

Sample #2 = 1.5720 (0.0100)

Sample #3 = 1.5630 (0.0210)

Sample #4 = 1.5470 (0.0440)

Aug % Abs = 1.5607 (0.0250)

STD DEV = 0.0127 (0.0173)

REL STD DEV = 0.811 (69.397)

Channel 2

Sample % Abs (% Abs Ref)

Sample #1 = 1.5650 (-0.0110)

Sample #2 = 1.5720 (0.0100)

Sample #3 = 1.5630 (0.0210)

Sample #4 = 1.5470 (0.0440)

Aug % Abs = 1.5607 (0.0250)

STD DEV = 0.0127 (0.0173)

REL STD DEV = 0.811 (69.397)

Channel 1

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.086

Std Dev = 0.01 Rel Std Dev = 12.25

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.863

Std Dev = 0.03 Rel Std Dev = 3.36

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 2.052

Std Dev = 0.02 Rel Std Dev = 1.18

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.934

Std Dev = 0.01 Rel Std Dev = 0.23

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.750

Std Dev = 0.00 Rel Std Dev = 0.04

Zero Order Coef = -186.55

First Order Coef = 2365.01

Second Order Coef = 26.32

Standard Deviation = 19.386513

Channel 2

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.082

Std Dev = 0.01 Rel Std Dev = 9.90

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.561

Std Dev = 0.01 Rel Std Dev = 0.81

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.773

Std Dev = 0.01 Rel Std Dev = 0.31

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 7.153

Std Dev = 0.01 Rel Std Dev = 0.11

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 10.464

Std Dev = 0.02 Rel Std Dev = 0.15

Zero Order Coef = -107.50

First Order Coef = 1262.16

Second Order Coef = 10.96

Standard Deviation = 34.125679

Channel 2

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.082

Std Dev = 0.01 Rel Std Dev = 9.90

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.561

Std Dev = 0.01 Rel Std Dev = 0.81

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.773

Std Dev = 0.01 Rel Std Dev = 0.31

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 7.153

Std Dev = 0.01 Rel Std Dev = 0.11

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 10.464

Std Dev = 0.02 Rel Std Dev = 0.15

Zero Order Coef = -107.50

First Order Coef = 1262.16

Second Order Coef = 10.96

Standard Deviation = 34.125679

Channel 2

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.082

Std Dev = 0.01 Rel Std Dev = 9.90

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.561

Std Dev = 0.01 Rel Std Dev = 0.81

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.773

Std Dev = 0.01 Rel Std Dev = 0.31

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 7.153

Std Dev = 0.01 Rel Std Dev = 0.11

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 10.464

Std Dev = 0.02 Rel Std Dev = 0.15

Zero Order Coef = -107.50

First Order Coef = 1262.16

Second Order Coef = 10.96

Standard Deviation = 34.125679

Channel 2

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.082

Std Dev = 0.01 Rel Std Dev = 9.90

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.561

Std Dev = 0.01 Rel Std Dev = 0.81

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.773

Std Dev = 0.01 Rel Std Dev = 0.31

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 7.153

Std Dev = 0.01 Rel Std Dev = 0.11

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 10.464

Std Dev = 0.02 Rel Std Dev = 0.15

Zero Order Coef = -107.50

First Order Coef = 1262.16

Second Order Coef = 10.96

Standard Deviation = 34.125679

Channel 2

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.082

Std Dev = 0.01 Rel Std Dev = 9.90

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.561

Std Dev = 0.01 Rel Std Dev = 0.81

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.773

Std Dev = 0.01 Rel Std Dev = 0.31

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 7.153

Std Dev = 0.01 Rel Std Dev = 0.11

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 10.464

Std Dev = 0.02 Rel Std Dev = 0.15

Zero Order Coef = -107.50

First Order Coef = 1262.16

Second Order Coef = 10.96

Standard Deviation = 34.125679

Solution Stats Quadratic Fit Chan 1

Act Fit Residual

g/210L g/210L g/210L

0.000 0.000 -0.0004

0.040 0.039 0.0007

0.100 0.100 -0.0003

0.200 0.200 -0.0000

0.300 0.300 0.0000

Solution Stats Quadratic Fit Chan 2

Act Fit Residual

g/210L g/210L g/210L

0.000 -0.000 0.0001

0.040 0.040 0.0003

0.100 0.101 -0.0010

0.200 0.199 0.0009

0.300 0.300 -0.0003

Sol Value = 0.080 g/210L \*\*\*

Fit value = 0.3810 mg/l %%%

Samples Taken = 4, Discarded = 1

Channel 1

Sample #1 = 3093.00

Sample #2 = 3148.00

Sample #3 = 3199.00

Sample #4 = 3140.00

Average Result = 3162.3333

STD DEV = 32.0052

REL STD DEV = 1.012

Channel 2

Sample #1 = 3360.00

Sample #2 = 3389.00

Sample #3 = 3402.00

Sample #4 = 3409.00

Average Result = 3400.0000

STD DEV = 10.1489

REL STD DEV = 0.298

Channel 2

Sample #1 = 3360.00

Sample #2 = 3389.00

Sample #3 = 3402.00

# Post Calibration Adjustment

Stability Checks #80-005113 Hillsborough County S.O. 5/17/17 *RMB*

*005*

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005113  
05/17/2017  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 12:02 |
| Control Test       | 0.049  | 12:02 |
| Air Blank          | 0.000  | 12:03 |
| Control Test       | 0.048  | 12:04 |
| Air Blank          | 0.000  | 12:04 |
| Control Test       | 0.049  | 12:05 |
| Air Blank          | 0.000  | 12:05 |
| Control Test Stats |        |       |
| Average            | 0.0487 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 1.1863 |       |

*RMB*

Operator's Signature

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005113  
05/17/2017  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 12:07 |
| Control Test       | 0.078  | 12:07 |
| Air Blank          | 0.000  | 12:08 |
| Control Test       | 0.078  | 12:08 |
| Air Blank          | 0.000  | 12:09 |
| Control Test       | 0.078  | 12:10 |
| Air Blank          | 0.000  | 12:10 |
| Control Test Stats |        |       |
| Average            | 0.0780 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

*RMB*

Operator's Signature

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005113  
05/17/2017  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 12:12 |
| Control Test       | 0.195  | 12:12 |
| Air Blank          | 0.000  | 12:13 |
| Control Test       | 0.195  | 12:14 |
| Air Blank          | 0.000  | 12:14 |
| Control Test       | 0.195  | 12:15 |
| Air Blank          | 0.000  | 12:16 |
| Control Test Stats |        |       |
| Average            | 0.1950 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

*RMB*

Operator's Signature

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005113  
05/17/2017  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 12:17 |
| Control Test       | 0.080  | 12:17 |
| Air Blank          | 0.000  | 12:18 |
| Control Test       | 0.081  | 12:18 |
| Air Blank          | 0.000  | 12:19 |
| Control Test       | 0.080  | 12:19 |
| Air Blank          | 0.000  | 12:20 |
| Control Test Stats |        |       |
| Average            | 0.0803 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 0.7187 |       |

*SP*

*RMB*

Operator's Signature

*SP*