



PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005337  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:30
Control Test	0.048	13:31
Air Blank	0.000	13:31
Control Test	0.048	13:32
Air Blank	0.000	13:32
Control Test	0.049	13:33
Air Blank	0.000	13:34
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

*P Murphy*  
 -----  
 Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005337  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:24
Control Test	0.078	13:25
Air Blank	0.000	13:26
Control Test	0.078	13:26
Air Blank	0.000	13:27
Control Test	0.078	13:28
Air Blank	0.000	13:28
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*P Murphy*  
 -----  
 Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005337  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:19
Control Test	0.199	13:19
Air Blank	0.000	13:20
Control Test	0.198	13:21
Air Blank	0.000	13:21
Control Test	0.198	13:22
Air Blank	0.000	13:23
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

*P Murphy*  
 -----  
 Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005337  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:01
Control Test	0.081	14:01
Air Blank	0.000	14:01
Control Test	0.080	14:02
Air Blank	0.000	14:02
Control Test	0.080	14:02
Air Blank	0.000	14:03
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

*P Murphy*  
 -----  
 Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005337  
 11/02/2017  
 Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
 1: Rate (Liters/min) = 5  
    SQRT(Diff) ) = 5.566  
 2: Rate (Liters/min) = 15  
    SQRT(Diff) ) = 10.953  
 3: Rate (Liters/min) = 30  
    SQRT(Diff) ) = 21.023  
 Dependent Data Scale Factor = 100000 L/min  
 Independent Data Scale Factor = 256  
 Rounded Slope = 625  
 Rounded Intercept = -336950  
 Correlation = 0.99831

*PJM*

11/6/17  
*JD*