

DAYTONA BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005537
08/31/2017
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:22 |
| Control Test | 0.047 | 13:22 |
| Air Blank | 0.000 | 13:23 |
| Control Test | 0.048 | 13:23 |
| Air Blank | 0.000 | 13:24 |
| Control Test | 0.049 | 13:25 |
| Air Blank | 0.000 | 13:25 |
| Control Test Stats | | |
| Average | 0.0480 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 2.0833 | |

P Murphy
Operator's Signature

DAYTONA BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005537
08/31/2017
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:32 |
| Control Test | 0.079 | 13:33 |
| Air Blank | 0.000 | 13:34 |
| Control Test | 0.079 | 13:34 |
| Air Blank | 0.000 | 13:35 |
| Control Test | 0.080 | 13:36 |
| Air Blank | 0.000 | 13:36 |
| Control Test Stats | | |
| Average | 0.0793 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7277 | |

P Murphy
Operator's Signature

DAYTONA BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005537
08/31/2017
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:27 |
| Control Test | 0.199 | 13:28 |
| Air Blank | 0.000 | 13:29 |
| Control Test | 0.198 | 13:29 |
| Air Blank | 0.000 | 13:30 |
| Control Test | 0.198 | 13:31 |
| Air Blank | 0.000 | 13:31 |
| Control Test Stats | | |
| Average | 0.1983 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.2911 | |

P Murphy
Operator's Signature

DAYTONA BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005537
08/31/2017
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 13:40 |
| Control Test | 0.080 | 13:40 |
| Air Blank | 0.000 | 13:40 |
| Control Test | 0.080 | 13:41 |
| Air Blank | 0.000 | 13:41 |
| Control Test | 0.080 | 13:42 |
| Air Blank | 0.000 | 13:42 |
| Control Test Stats | | |
| Average | 0.0800 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

D63

P Murphy
Operator's Signature

DS

9/13/17
JW