

INSTRUMENT PROCESSING SHEET

Agency Miami Dade Police Department S/N 80-005550
 Date In 1/10/2017 Date Out 1/10/2017 Ship P/U H/D CMI EE

Intake Performed By DELL

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other ANTI STATIC BAG

Notes: _____

Quality Checks Performed By DELL

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 183
 Flow Verification (L/s)
 Flow Column # ATP101
 32mm 164 (.139 - .169)
 36mm 183 (.156 - .190)
 53mm 250 (.228 - .278)
 103mm 507 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28663

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>503967</u>	<u>201507A</u> <u>07/14/2017</u>
0.08	<u>503968</u>	<u>201601F</u> <u>01/26/2018</u>
0.20	<u>503969</u>	<u>201604C</u> <u>04/06/2018</u>
0.08 DGS	N/A	<u>A6605301</u> <u>02/22/2018</u>

Flow Calibration Performed By DELL

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 JAN 18 2017
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DELL

Barometric Pressure 1028 Gauge
 ID# 68639 1027 Instrument

Mouth Alcohol Solution Lot # 2016A
 Acetone Stock Solution Lot # 2016B

Simulator	Serial Number
0.00	<u>SD 3965</u>
Interferent	<u>SD 3966</u>
0.05	<u>SD 3967</u>
0.08	<u>SD 3968</u>
0.20	<u>SD 3969</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: **E-MAILED** **APPROVED**
1/10/2017
DA/AC OK 1/17/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Wickland
 Quality Control Review

4/18/17
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-005550	Miami Dade Police Department	01/10/2017	<i>lll</i>

0.05g/210L 0.047 to 0.053 0.08g/210L 0.077 to 0.083 0.20g/210L 0.194 to 0.206 DGS 0.08g/210L 0.077 to 0.083

MIAMI DADE PD
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-005550
01/10/2017
Software: 8100.27

Test g/210L Time

Air Blank 0.000 10:38
Control Test 0.049 10:39
Air Blank 0.000 10:39
Control Test 0.049 10:40
Air Blank 0.000 10:40
Control Test 0.049 10:41
Air Blank 0.000 10:42
Control Test Stats
Average 0.0487
Std Dev 0.0006
Rel Std Dev(%) 1.1963

MIAMI DADE PD
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-005550
01/10/2017
Software: 8100.27

Test g/210L Time

Air Blank 0.000 10:43
Control Test 0.078 10:43
Air Blank 0.000 10:44
Control Test 0.079 10:45
Air Blank 0.000 10:45
Control Test 0.079 10:46
Air Blank 0.000 10:46
Control Test Stats
Average 0.0787
Std Dev 0.0006
Rel Std Dev(%) 0.7339

MIAMI DADE PD
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-005550
01/10/2017
Software: 8100.27

Test g/210L Time

Air Blank 0.000 10:48
Control Test 0.195 10:48
Air Blank 0.000 10:49
Control Test 0.196 10:49
Air Blank 0.000 10:50
Control Test 0.197 10:51
Air Blank 0.000 10:51
Control Test Stats
Average 0.1960
Std Dev 0.0010
Rel Std Dev(%) 0.5102

MIAMI DADE PD
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-005550
01/10/2017
Software: 8100.27

Test g/210L Time

Air Blank 0.000 10:53
Control Test 0.078 10:53
Air Blank 0.000 10:53
Control Test 0.079 10:54
Air Blank 0.000 10:54
Control Test 0.079 10:55
Air Blank 0.000 10:55
Control Test Stats
Average 0.0787
Std Dev 0.0006
Rel Std Dev(%) 0.7339

lll
Operator's Signature

lll
Operator's Signature

lll
Operator's Signature

lll
Operator's Signature

gpm
sk