

INSTRUMENT PROCESSING SHEET

Agency FHP - Tampa

S/N 30-006046

Date In 5/3/17

Date Out 5/4/17

Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 216
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm 0.152 (.139 - .169)
 36mm 0.164 (.156 - .190)
 53mm 0.230 (.228 - .278)
 103mm 0.503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD3964</u>	<u>201611B</u> <u>11/5/18</u>
0.20	<u>G11621</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG026605</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 47mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAY 04 2017
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1007 Gauge
 ID# 28421 1007 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2879</u>
Interferent	<u>G8144</u>
0.05	<u>DR2035</u>
0.08	<u>SD3964</u>
0.20	<u>G11621</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/QC OK SP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

[Signature]
 Quality Control Review

5/4/17
 Date

Stability Checks # 80-006046 FL Highway Patrol 5/4/17 ~~RMS~~

SUS

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006046
05/04/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:16
Control Test	0.049	09:17
Air Blank	0.000	09:17
Control Test	0.049	09:18
Air Blank	0.000	09:18
Control Test	0.049	09:19
Air Blank	0.000	09:20
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

~~RMS~~

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006046
05/04/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:20
Control Test	0.081	09:21
Air Blank	0.000	09:22
Control Test	0.081	09:22
Air Blank	0.000	09:23
Control Test	0.081	09:24
Air Blank	0.000	09:24
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

~~RMS~~

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006046
05/04/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:25
Control Test	0.201	09:26
Air Blank	0.000	09:27
Control Test	0.199	09:27
Air Blank	0.000	09:28
Control Test	0.199	09:29
Air Blank	0.000	09:29
Control Test Stats		
Average	0.1997	
Std Dev	0.0012	
Rel. Std Dev(%)	0.5783	

~~RMS~~

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006046
05/04/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:12
Control Test	0.079	09:12
Air Blank	0.000	09:13
Control Test	0.079	09:13
Air Blank	0.000	09:14
Control Test	0.079	09:14
Air Blank	0.000	09:14
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

SP

312

~~RMS~~

Operator's Signature