

## INSTRUMENT PROCESSING SHEET

Agency Ocala PD S/N 80-006214  
 Date In 6/6/17 Date Out 6/21/17  Ship  P/U  H/D  CMI  EE

**Intake** Performed By SP

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case  Handle  
 Dry Gas Holder  Feet  
 Keyboard/Plug  Back/Plugs  
 Screws tight  Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By SP

Lab Temp °C 20.6  
 Breath Tube Screen  
 Replace External O-Rings  
 Instrument Set Up Verified  
 R-Value 207  
 Flow Verification (L/s)  
 Flow Column # ATP105  
 32mm 156 (.139 - .169)  
 36mm 175 (.156 - .190)  
 53mm 246 (.228 - .278)  
 103mm 515 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	201603D 3/18/18
0.08	G8149	201611B 11/15/18
0.20	G11621	201604E 4/15/18
0.08 DGS	N/A	AG-621604 9/22/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Quality Checks Cont.** Performed By SP

**Simulator Temperatures °C**  
 External Digital Therm. ID#: 300504  
 34°C +/- .2 Serial #: G11739  
 34°C +/- .2 Serial #: G8149  
 34°C +/- .2 Serial #: G11621

RECEIVED  
 JUN 21 2017  
 FDLE  
 Alcohol Testing Program

**Calibration Adjustment** Performed By \_\_\_\_\_

Calibration Adjustment N/A  
 Calibration Adjustment Complete

Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By SP

Barometric Pressure 1013 Gauge  
 ID# 26932 1014 Instrument

Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G11739
0.08	G8149
0.20	G11621

**Attachments**

Form 41  Calibration Adjustment  
 Pre-Stability Tests  Post-Stability Tests  
 Flow Calibration  Other \_\_\_\_\_

Notes/Suggested Service:  
QA/OC OK @ 6/21/2017

J. Graham 6/21/17  
 Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

80-006214  
 Stability checks  
 6/21/17

INTOXILYZER 8000  
 Instrument Initialization  
 06:32 06/21/2017

OCALA POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-006214  
 06/21/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:16
Control Test	0.079	08:17
Air Blank	0.000	08:17
Control Test	0.079	08:18
Air Blank	0.000	08:18
Control Test	0.079	08:18
Air Blank	0.000	08:19
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Operator's Signature

OCALA POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-006214  
 06/21/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:20
Control Test	0.078	08:21
Air Blank	0.000	08:21
Control Test	0.078	08:22
Air Blank	0.000	08:23
Control Test	0.078	08:23
Air Blank	0.000	08:24
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

OCALA POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-006214  
 06/21/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:25
Control Test	0.048	08:26
Air Blank	0.000	08:26
Control Test	0.048	08:27
Air Blank	0.000	08:28
Control Test	0.048	08:28
Air Blank	0.000	08:29
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

OCALA POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-006214  
 06/21/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:30
Control Test	0.196	08:31
Air Blank	0.000	08:31
Control Test	0.196	08:32
Air Blank	0.000	08:32
Control Test	0.196	08:33
Air Blank	0.000	08:34
Control Test Stats		
Average	0.1960	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Boon

Operator's Signature

JD