



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Lake City

S/N 80-006221

Date In 1/11/17

Date Out 1/12/17

Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
ok Case ok Handle  
ok Dry Gas Holder ok Feet  
ok Keyboard/Plug ok Back/Plugs  
ok Screws tight ok Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 166  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm 148 (.139 - .169)  
 36mm 166 (.156 - .190)  
 53mm 251 (.228 - .278)  
 103mm 497 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 26932

Stability Checks

| Simulator | Serial # | Lot #/Exp           |
|-----------|----------|---------------------|
| 0.05      | DR2035   | 201603D<br>3/8/18   |
| 0.08      | SD1011   | 201601F<br>1/26/18  |
| 0.20      | SD1025   | 201601C<br>4/5/18   |
| 0.08 DGS  | N/A      | A1626605<br>9/22/18 |

**Flow Calibration** Performed By [Signature]

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 47mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
JAN 12 2017  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete  
 Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000     |               | N/A        | N/A        |
| 0.040     |               |            |            |
| 0.100     |               |            |            |
| 0.200     |               |            |            |
| 0.400     |               |            |            |
| 0.080 DGS | N/A           |            |            |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05      |               |            |            |
| 0.08      |               |            |            |
| 0.20      |               |            |            |
| 0.08 DGS  | N/A           |            |            |

**Department Inspection** Performed By [Signature]

Barometric Pressure 1025 Gauge  
 ID# 26932 1025 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

| Simulator   | Serial Number |
|-------------|---------------|
| 0.00        | SD1019        |
| Interferent | SD1021        |
| 0.05        | DR2035        |
| 0.08        | SD1011        |
| 0.20        | SD1025        |

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: QC: MBS

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kurland  
Quality Control Review

1/12/17  
Date

80-006221  
 Stud. 1117 (lots)  
 1/12/17

INTOXILYZER 8000  
 Instrument Initialization  
 10:05 01/12/2017

LAKE CITY PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-006221  
 01/12/2017  
 Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 08:36 |
| Control Test       | 0.080  | 08:37 |
| Air Blank          | 0.000  | 08:37 |
| Control Test       | 0.080  | 08:38 |
| Air Blank          | 0.000  | 08:38 |
| Control Test       | 0.080  | 08:38 |
| Air Blank          | 0.000  | 08:39 |
| Control Test Stats |        |       |
| Average            | 0.0800 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

DS

Operator's Signature

LAKE CITY PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-006221  
 01/12/2017  
 Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 08:41 |
| Control Test       | 0.080  | 08:42 |
| Air Blank          | 0.000  | 08:42 |
| Control Test       | 0.080  | 08:42 |
| Air Blank          | 0.000  | 08:43 |
| Control Test       | 0.080  | 08:43 |
| Air Blank          | 0.000  | 08:44 |
| Control Test Stats |        |       |
| Average            | 0.0800 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

Operator's Signature

DS

LAKE CITY PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-006221  
 01/12/2017  
 Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 08:47 |
| Control Test       | 0.049  | 08:48 |
| Air Blank          | 0.000  | 08:48 |
| Control Test       | 0.049  | 08:49 |
| Air Blank          | 0.000  | 08:50 |
| Control Test       | 0.049  | 08:50 |
| Air Blank          | 0.000  | 08:51 |
| Control Test Stats |        |       |
| Average            | 0.0490 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

Operator's Signature

LAKE CITY PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-006221  
 01/12/2017  
 Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 08:57 |
| Control Test       | 0.200  | 08:58 |
| Air Blank          | 0.000  | 08:58 |
| Control Test       | 0.200  | 08:59 |
| Air Blank          | 0.000  | 08:59 |
| Control Test       | 0.200  | 08:59 |
| Air Blank          | 0.000  | 09:00 |
| Control Test       | 0.200  | 09:01 |
| Air Blank          | 0.000  | 09:01 |
| Control Test Stats |        |       |
| Average            | 0.2000 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

Operator's Signature

RK