

## INSTRUMENT PROCESSING SHEET

RECEIVED

MAY 01 2017

Agency Escombia County S/N 80-006435  
 Date In 4/20/17 Date Out 4/27/17

Ship  P/U  H/D  CMI  EDLE  
 Alcohol Testing Program

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case OK Handle OK  
 Dry Gas Holder OK Feet OK  
 Keyboard/Plug OK Back/Plugs  
 Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 195  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm 164 (.139 - .169)  
 36mm 167 (.156 - .190)  
 53mm 234 (.228 - .278)  
 103mm 500 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	DR2035	201603D 3/8/18
0.08	SD3964	201611B 11/15/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG626604 9/22/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1009 Gauge  
 ID# 28421 1007 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	DR2035
0.08	SD3964
0.20	SD1025

**Attachments**

Form 41  Optical Bench Cal  
 Pre-Stability Tests  Post-Stability Tests  
 Flow Calibration  Other \_\_\_\_\_

Notes: OC [Signature] 4/28/17

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

[Signature]  
 Quality Control Review

5/1/17  
 Date

ESCAMBIA COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006435  
04/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:34
Control Test	0.048	08:34
Air Blank	0.000	08:35
Control Test	0.049	08:35
Air Blank	0.000	08:36
Control Test	0.049	08:37
Air Blank	0.000	08:37
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

*P Murphy*  
-----  
Operator's Signature

ESCAMBIA COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006435  
04/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:45
Control Test	0.080	08:46
Air Blank	0.000	08:46
Control Test	0.079	08:47
Air Blank	0.000	08:47
Control Test	0.080	08:48
Air Blank	0.000	08:49
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

*P Murphy*  
-----  
Operator's Signature

ESCAMBIA COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006435  
04/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:40
Control Test	0.198	08:40
Air Blank	0.000	08:41
Control Test	0.198	08:41
Air Blank	0.000	08:42
Control Test	0.198	08:43
Air Blank	0.000	08:43
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*P Murphy*  
-----  
Operator's Signature

ESCAMBIA COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006435  
04/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.081	08:51
Air Blank	0.000	08:52
Control Test	0.082	08:52
Air Blank	0.000	08:53
Control Test	0.081	08:53
Air Blank	0.000	08:54
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

DGS

*P Murphy*  
-----  
Operator's Signature