

INSTRUMENT PROCESSING SHEET

Agency Eschambia County S/N 80-006436

Date In 3/23/17

Date Out 3/29/17

Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 229
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .164 (.139 - .169)
 36mm .175 (.156 - .190)
 53mm .253 (.228 - .278)
 103mm .492 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Simulator	Serial #	Lot #/Exp
0.05	DR2035	201603D 3/8/18
0.08	503964	201601F 1/26/18
0.20	501025	201604C 4/5/18
0.08 DGS	N/A	AG626604 9/22/18

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1014 Gauge
 ID# 28421 1015 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	501019
Interferent	501021
0.05	DR2035
0.08	503964
0.20	501025

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: [Signature] 3/29/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Butt Kuiteland

Quality Control Review

3/29/17
Date

RECEIVED
 APR 03 2017
 FDLE
 Alcohol Testing Program

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006436
03/29/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:36
Control Test	0.048	07:36
Air Blank	0.000	07:37
Control Test	0.049	07:38
Air Blank	0.000	07:38
Control Test	0.049	07:39
Air Blank	0.000	07:39
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	



Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006436
03/29/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:41
Control Test	0.079	07:42
Air Blank	0.000	07:42
Control Test	0.080	07:43
Air Blank	0.000	07:43
Control Test	0.080	07:44
Air Blank	0.000	07:45
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	



Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006436
03/29/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:46
Control Test	0.199	07:47
Air Blank	0.000	07:47
Control Test	0.200	07:48
Air Blank	0.000	07:49
Control Test	0.200	07:49
Air Blank	0.000	07:50
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	



Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006436
03/29/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:51
Control Test	0.078	07:52
Air Blank	0.000	07:52
Control Test	0.078	07:52
Air Blank	0.000	07:53
Control Test	0.079	07:53
Air Blank	0.000	07:54
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	



D65



Operator's Signature