

INSTRUMENT PROCESSING SHEET

Agency Palm Beach CSO S/N 80-006476
 Date In 2/6/17 Date Out 2/6/17 Ship P/U H/D CMI EE

Intake Performed By SP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 183
 Flow Verification (L/s)
 Flow Column # ATP106
 32mm .152 (.139 - .169)
 36mm .171 (.156 - .190)
 53mm .238 (.228 - .278)
 103mm .500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 286603

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|--------------------|
| 0.05 | SD1014 | 201507A 7-14-17 |
| 0.08 | SD1015 | 201601F 1-26-18 |
| 0.20 | SD1017 | 201505A 5-12-17 |
| 0.08 DGS | N/A | AG60530 2-22-18 |

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 FEB 10 2017
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.400 | | | |
| 0.080 DGS | N/A | | |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Department Inspection Performed By SP

Barometric Pressure 1020 Gauge
 ID# 68639 1020 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | G8147 |
| Interferent | G12100 |
| 0.05 | SD1014 |
| 0.08 | SD1015 |
| 0.20 | SD1017 |

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: **E-MAILED**

QA/QC OK BEAR

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Whitland
 Quality Control Review

2/10/17
 Date

| TYPE OF TEST | SERIAL NUMBER | AGENCY | DATE | PERFORMED BY |
|--------------|---------------|----------------|--------|--------------|
| Stabilities | 80-006476 | Palm Beach CSO | 2/6/17 | SJP |

57

0.05g/210L

0.08g/210L

0.20g/210L

DGS 0.08g/210L

PALM BEACH SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006476
02/06/2017
Software: 8100.27

PALM BEACH SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006476
02/06/2017
Software: 8100.27

PALM BEACH SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006476
02/06/2017
Software: 8100.27

PALM BEACH SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006476
02/06/2017
Software: 8100.27

Test g/210L Time

Air Blank 0.000 09:24
Control Test 0.049 09:25
Air Blank 0.000 09:25
Control Test 0.049 09:26
Air Blank 0.000 09:27
Control Test 0.049 09:27
Air Blank 0.000 09:28

Control Test Stats
Average 0.0490
Std Dev 0.0000
Rel Std Dev(%) 0.0000

Test g/210L Time

Air Blank 0.000 09:18
Control Test 0.079 09:19
Air Blank 0.000 09:20
Control Test 0.079 09:20
Air Blank 0.000 09:21
Control Test 0.080 09:22
Air Blank 0.000 09:22

Control Test Stats
Average 0.0793
Std Dev 0.0006
Rel Std Dev(%) 0.7277

Test g/210L Time

Air Blank 0.000 09:13
Control Test 0.201 09:14
Air Blank 0.000 09:14
Control Test 0.201 09:15
Air Blank 0.000 09:15
Control Test 0.202 09:16
Air Blank 0.000 09:17

Control Test Stats
Average 0.2013
Std Dev 0.0006
Rel Std Dev(%) 0.2868

Test g/210L Time

Air Blank 0.000 09:31
Control Test 0.081 09:32
Air Blank 0.000 09:32
Control Test 0.080 09:33
Air Blank 0.000 09:33
Control Test 0.080 09:34
Air Blank 0.000 09:34

Control Test Stats
Average 0.0803
Std Dev 0.0006
Rel Std Dev(%) 0.7187

DGS

SJP
Operator's Signature

SJP
Operator's Signature

SJP
Operator's Signature

SJP
Operator's Signature