

INSTRUMENT PROCESSING SHEET

Agency Manatee County S/N 80-006694
 Date In 7/18/17 Date Out 7/20/17 Ship P/U H/D CMI EE

RECEIVED
 JUL 20 2017
 FDLE
 Alcohol Testing Program

Intake Performed By SP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Lab Temp °C 23.9
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 199
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm .148 (.139 - .169)
 36mm .171 (.156 - .190)
 53mm .246 (.228 - .278)
 103mm .519 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>SD3962</u>	<u>2016030</u> <u>3-8-18</u>
0.08	<u>SD1013</u>	<u>2016118</u> <u>11-15-18</u>
0.20	<u>SD3933</u>	<u>201702B</u> <u>2-23-19</u>
0.08 DGS	N/A	<u>AG624605</u> <u>9-22-18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By SP

Simulator Temperatures °C
 External Digital Therm. ID#: 300504
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD1013
 34°C +/- .2 Serial #: SD3933

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure
 ID# 28427 Gauge 1014
1015 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>SD3962</u>
0.08	<u>SD1013</u>
0.20	<u>SD3933</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
QA/QC on 7/20/17

JJ Graham 7/20/17
 Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

MANATEE COUNTY SO - #80-006694 - STABILITY CHECKS

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006694
 07/20/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:48
Control Test	0.048	08:49
Air Blank	0.000	08:49
Control Test	0.048	08:50
Air Blank	0.000	08:51
Control Test	0.048	08:51
Air Blank	0.000	08:52
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

JD

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006694
 07/20/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:44
Control Test	0.077	08:44
Air Blank	0.000	08:45
Control Test	0.078	08:46
Air Blank	0.000	08:46
Control Test	0.078	08:47
Air Blank	0.000	08:47
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

SP

Operator's Signature

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006694
 07/20/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:01
Control Test	0.080	09:01
Air Blank	0.000	09:02
Control Test	0.079	09:02
Air Blank	0.000	09:02
Control Test	0.079	09:03
Air Blank	0.000	09:03
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

SP

Operator's Signature

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006694
 07/20/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:04
Control Test	0.199	09:05
Air Blank	0.000	09:05
Control Test	0.199	09:06
Air Blank	0.000	09:06
Control Test	0.199	09:07
Air Blank	0.000	09:08
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

RS