

## INSTRUMENT PROCESSING SHEET

Agency FHP S/N 80-006760  
 Date In 3/1/17 Date Out 3/7/2017  
 Ship  P/U  H/D  CMI  EE

**Intake** Performed By SP

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case  Handle  
 Dry Gas Holder  Feet  
 Keyboard/Plug  Back/Plugs  
 Screws tight  Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other \_\_\_\_\_

Notes: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Quality Checks** Performed By pgm

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 236  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm .160 (.139 - .169)  
 36mm .167 (.156 - .190)  
 53mm .234 (.228 - .278)  
 103mm .492 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>611739</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>68149</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>611621</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>80626605</u> <u>9/22/18</u>

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**  
 \_\_\_\_\_  
 \_\_\_\_\_

RECEIVED  
 MAR 20 2017  
 FDLE  
 Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By pgm

Barometric Pressure 1024 Gauge  
 ID# 26932 1024 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>62879</u>
Interferent	<u>68144</u>
0.05	<u>611739</u>
0.08	<u>68149</u>
0.20	<u>611621</u>

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: QC SP 3/20/17  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kurland  
 Quality Control Review

3/20/17  
 Date

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006760  
03/07/2017  
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	10:32	
Control Test	0.048	10:33	
Air Blank	0.000	10:34	
Control Test	0.049	10:34	
Air Blank	0.000	10:35	
Control Test	0.048	10:36	
Air Blank	0.000	10:36	
Control Test Stats			
Average	0.0483		
Std Dev	0.0006		
Rel Std Dev(%)	1.1945		

*P Murphy*  
Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006760  
03/07/2017  
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	10:42	
Control Test	0.078	10:43	
Air Blank	0.000	10:43	
Control Test	0.077	10:44	
Air Blank	0.000	10:45	
Control Test	0.077	10:45	
Air Blank	0.000	10:46	
Control Test Stats			
Average	0.0773		
Std Dev	0.0006		
Rel Std Dev(%)	0.7466		

*P Murphy*  
Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006760  
03/07/2017  
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	10:37	
Control Test	0.195	10:38	
Air Blank	0.000	10:38	
Control Test	0.196	10:39	
Air Blank	0.000	10:40	
Control Test	0.196	10:40	
Air Blank	0.000	10:41	
Control Test Stats			
Average	0.1957		
Std Dev	0.0006		
Rel Std Dev(%)	0.2951		

*P Murphy*  
Operator's Signature

BK

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006760  
03/07/2017  
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	10:51	
Control Test	0.080	10:51	
Air Blank	0.000	10:51	
Control Test	0.080	10:52	
Air Blank	0.000	10:52	
Control Test	0.080	10:53	
Air Blank	0.000	10:53	
Control Test Stats			
Average	0.0800		
Std Dev	0.0000		
Rel Std Dev(%)	0.0000		

*GA*

DGS

*P Murphy*  
Operator's Signature