



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP

S/N 80-006770

Date In 3/2/17

Date Out 3/9/17

Ship P/U H/D CMI EE

Intake Performed By DD

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 245
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .148 (.139 - .169)
 36mm .164 (.156 - .190)
 53mm .230 (.228 - .278)
 103mm .490 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G3709</u>	<u>201003D</u> <u>3-8-18</u>
0.08	<u>SD3964</u>	<u>201001F</u> <u>1-20-18</u>
0.20	<u>DR3856</u>	<u>201004C</u> <u>4-5-18</u>
0.08 DGS	N/A	<u>AG1026005</u> <u>9-22-18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 19mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
MAR 13 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure
 ID# 28427 1020 Gauge
1020 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>G3709</u>
0.08	<u>SD3964</u>
0.20	<u>DR3856</u>

Attachments

Form 41 X2
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: OK 3/9/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kristofand

Quality Control Review

3/13/17

Date

STABILITY CHECKS INSTRUMENT # 80-006770

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006770
 03/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:29
Control Test	0.048	09:29
Air Blank	0.000	09:30
Control Test	0.048	09:31
Air Blank	0.000	09:31
Control Test	0.048	09:32
Air Blank	0.000	09:32
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

 Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006770
 03/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:24
Control Test	0.077	09:25
Air Blank	0.000	09:25
Control Test	0.077	09:26
Air Blank	0.000	09:26
Control Test	0.078	09:27
Air Blank	0.000	09:28
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

SP

 Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006770
 03/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:40
Control Test	0.195	09:41
Air Blank	0.000	09:42
Control Test	0.195	09:42
Air Blank	0.000	09:43
Control Test	0.196	09:44
Air Blank	0.000	09:44
Control Test Stats		
Average	0.1953	
Std Dev	0.0006	
Rel Std Dev(%)	0.2956	

SP

 Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006770
 03/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.079	09:46
Air Blank	0.000	09:47
Control Test	0.079	09:47
Air Blank	0.000	09:47
Control Test	0.079	09:48
Air Blank	0.000	09:48
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

 Operator's Signature

DEIS

[Handwritten mark]

BK