

INSTRUMENT PROCESSING SHEET

Agency FHP - Troop F

S/N 80-006771

Date In 4/20/17

Date Out 4/26/17

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake ☒ Registration ☒ Annual ☒ Return from CMI ☐ Return from Enforcement ☐ Electronics ☐ Other _____ Visual Inspection: ☒ Case ☒ Handle ☒ Dry Gas Holder ☒ Feet ☒ Keyboard/Plug ☒ Back/Plugs ☒ Screws tight ☒ Breath Hose Other Equipment: ☒ Power cord ☒ Printer Cable ☒ Other <u>Static Bag</u> Notes: 	Quality Checks Performed By <u>SP</u> ☒ Breath Tube Screen ☒ Replace O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>236</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>1148</u> (.139 - .169) 36mm <u>164</u> (.156 - .190) 53mm <u>230</u> (.228 - .278) 103mm <u>488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>G3709</u></td> <td><u>201603D</u> <u>3-8-18</u></td> </tr> <tr> <td>0.08</td> <td><u>DR1279</u></td> <td><u>201611B</u> <u>11-15-18</u></td> </tr> <tr> <td>0.20</td> <td><u>DR3856</u></td> <td><u>201604C</u> <u>4-5-18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG626605</u> <u>9-22-18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>G3709</u>	<u>201603D</u> <u>3-8-18</u>	0.08	<u>DR1279</u>	<u>201611B</u> <u>11-15-18</u>	0.20	<u>DR3856</u>	<u>201604C</u> <u>4-5-18</u>	0.08 DGS	N/A	<u>AG626605</u> <u>9-22-18</u>	Flow Calibration Performed By <u>SP</u> ☒ Flow Calibration N/A ☐ Flow Calibration Complete Flow Column # _____ ☐ 5L/min - 10mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Suggested Service
Simulator	Serial #	Lot #/Exp															
0.05	<u>G3709</u>	<u>201603D</u> <u>3-8-18</u>															
0.08	<u>DR1279</u>	<u>201611B</u> <u>11-15-18</u>															
0.20	<u>DR3856</u>	<u>201604C</u> <u>4-5-18</u>															
0.08 DGS	N/A	<u>AG626605</u> <u>9-22-18</u>															

Optical Bench Calibration Performed By _____
☒ Optical Bench Calibration N/A
☐ Optical Bench Calibration Complete
 Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Notes: Please change level 2 password to something unique.

QA/OC OK 4/26/17

Department Inspection Performed By SP
☒ Barometric Pressure 1009 Gauge
 ID# 28427 1009 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>G3709</u>
0.08	<u>DR1279</u>
0.20	<u>DR3856</u>

Attachments

- | | |
|---|---|
| ☒ Form 41 | ☐ Optical Bench Cal |
| ☒ Pre-Stability Tests | ☐ Post-Stability Tests |
| ☐ Flow Calibration | ☐ Other _____ |

- ☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC
☒ Return to/Place Into Evidentiary Use
☐ Remain Out of Evidentiary Use
☒ Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

STABILITY CHECKS- INSTRUMENT# 80-006771 - FHP

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006771
04/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.048	09:32
Air Blank	0.000	09:33
Control Test	0.049	09:33
Air Blank	0.000	09:34
Control Test	0.048	09:35
Air Blank	0.000	09:35
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006771
04/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:45
Control Test	0.077	09:46
Air Blank	0.000	09:47
Control Test	0.077	09:47
Air Blank	0.000	09:48
Control Test	0.078	09:49
Air Blank	0.000	09:49
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006771
04/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:40
Control Test	0.196	09:41
Air Blank	0.000	09:41
Control Test	0.195	09:42
Air Blank	0.000	09:42
Control Test	0.196	09:43
Air Blank	0.000	09:44
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006771
04/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:22
Control Test	0.080	09:23
Air Blank	0.000	09:23
Control Test	0.080	09:24
Air Blank	0.000	09:24
Control Test	0.080	09:24
Air Blank	0.000	09:25
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DG8

SP

SP

Operator's Signature

SP



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP

S/N 80-006771

Date In 3/2/17

Date Out 3/7/17

☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake ☒ Registration ☒ Annual ☐ Return from CMI ☐ Return from Enforcement ☐ Electronics ☐ Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: ☐ Power cord ☐ Printer Cable ☐ Other _____ Notes: _____ _____ _____ _____	Quality Checks Performed By <u>SP</u> ☒ Breath Tube Screen ☒ Replace O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>220</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>.148</u> (.139 - .169) 36mm <u>.171</u> (.156 - .190) 53mm <u>.234</u> (.228 - .278) 103mm <u>.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>DR2035</u></td> <td><u>201603D</u> <u>3-8-18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD1011</u></td> <td><u>201601F</u> <u>1-26-18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> <td><u>201604C</u> <u>4-5-18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG1026904</u> <u>9-22-18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>DR2035</u>	<u>201603D</u> <u>3-8-18</u>	0.08	<u>SD1011</u>	<u>201601F</u> <u>1-26-18</u>	0.20	<u>SD1025</u>	<u>201604C</u> <u>4-5-18</u>	0.08 DGS	N/A	<u>AG1026904</u> <u>9-22-18</u>	Flow Calibration Performed By _____ ☒ Flow Calibration N/A ☐ Flow Calibration Complete Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Suggested Service _____ _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	<u>DR2035</u>	<u>201603D</u> <u>3-8-18</u>															
0.08	<u>SD1011</u>	<u>201601F</u> <u>1-26-18</u>															
0.20	<u>SD1025</u>	<u>201604C</u> <u>4-5-18</u>															
0.08 DGS	N/A	<u>AG1026904</u> <u>9-22-18</u>															

RECEIVED
MAY 01 2017
FL
Alcohol Testing Program

Optical Bench Calibration Performed By _____ ☐ Optical Bench Calibration N/A ☐ Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																																
0.000		N/A	N/A																																																
0.040																																																			
0.100																																																			
0.200																																																			
0.400																																																			
0.080 DGS	N/A																																																		
Simulator	Serial Number	Lot Number	Expiration																																																
0.05																																																			
0.08																																																			
0.20																																																			
0.08 DGS	N/A																																																		

Notes: _____

Department Inspection Performed By _____ ☐ Barometric Pressure _____ Gauge ID# _____ Instrument Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.05</td> <td></td> </tr> <tr> <td>0.08</td> <td></td> </tr> <tr> <td>0.20</td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	0.00		Interferent		0.05		0.08		0.20	
Simulator	Serial Number												
0.00													
Interferent													
0.05													
0.08													
0.20													
Attachments ☐ Form 41 ☐ Pre-Stability Tests ☐ Flow Calibration ☐ Optical Bench Cal ☐ Post-Stability Tests ☐ Other _____													

☐ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC
☐ Return to/Place into Evidentiary Use
☐ Remain Out of Evidentiary Use
☐ Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

Revised November 2012

50

STABILITY CHECKS - INSTRUMENT # 80-006771

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006771
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:20
Control Test	0.047	12:20
Air Blank	0.000	12:21
Control Test	0.047	12:21
Air Blank	0.000	12:22
Control Test	0.048	12:23
Air Blank	0.000	12:23
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006771
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:13
Control Test	0.076	12:14
Air Blank	0.000	12:15
Control Test	0.077	12:15
Air Blank	0.000	12:16
Control Test	0.078	12:16
Air Blank	0.000	12:17
Control Test Stats		
Average	0.0770	
Std Dev	0.0010	
Rel Std Dev(%)	1.2987	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006771
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:27
Control Test	0.192	12:28
Air Blank	0.000	12:29
Control Test	0.194	12:29
Air Blank	0.000	12:30
Control Test	0.195	12:30
Air Blank	0.000	12:31
Control Test Stats		
Average	0.1937	
Std Dev	0.0015	
Rel Std Dev(%)	0.7887	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006771
03/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:24
Control Test	0.081	12:24
Air Blank	0.000	12:25
Control Test	0.081	12:25
Air Blank	0.000	12:26
Control Test	0.081	12:26
Air Blank	0.000	12:27
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP

Operator's Signature

SP