

INSTRUMENT PROCESSING SHEET

Agency Tallahassee PD S/N 80-006926
 Date In 10/11/17 Date Out 10/13/17 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input checked="" type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.2</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>168</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>.164</u> (.139 - .169) 36mm <u>.189</u> (.156 - .190) 53mm <u>.265</u> (.228 - .278) 103mm <u>.656</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>62835</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.08</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.20</td> <td><u>501025</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG702807</u> <u>3/29/19</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>	0.08	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.20	<u>501025</u>	<u>201707C</u> <u>7/24/19</u>	0.08 DGS	N/A	<u>AG702807</u> <u>3/29/19</u>	Flow Calibration Performed By <u>[Signature]</u> <input type="checkbox"/> Flow Calibration N/A <input checked="" type="checkbox"/> Flow Calibration Complete Flow Column # <u>ATP103</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>168</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>.152</u> (.139 - .169) 36mm <u>.164</u> (.156 - .190) 53mm <u>.230</u> (.228 - .278) 103mm <u>.503</u> (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>															
0.08	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>															
0.20	<u>501025</u>	<u>201707C</u> <u>7/24/19</u>															
0.08 DGS	N/A	<u>AG702807</u> <u>3/29/19</u>															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Quality Checks Cont. Performed By <u>[Signature]</u> Simulator Temperatures °C External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>62835</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>501025</u>															

RECEIVED
 OCT 17 2017
 FDLE
 Alcohol Testing Program

Calibration Adjustment Performed By _____																							
<input checked="" type="checkbox"/> Calibration Adjustment N/A <input type="checkbox"/> Calibration Adjustment Complete Barometric Pressure Gauge _____ ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.300																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

Department Inspection Performed By <u>[Signature]</u>	
<input checked="" type="checkbox"/> Barometric Pressure ID# <u>28421</u> Gauge <u>1013</u> Instrument <u>1013</u>	
Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u>	
Simulator	Serial Number
0.00	<u>62880</u>
Interferent	<u>62840</u>
0.05	<u>62835</u>
0.08	<u>DR1279</u>
0.20	<u>501025</u>
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	

Notes/Suggested Service:
QA/QC OK SP

[Signature] 10/17/17
 Quality Control Review Date

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

TALLAHASSEE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006926
 10/13/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:35
Control Test	0.048	08:35
Air Blank	0.000	08:36
Control Test	0.049	08:37
Air Blank	0.000	08:37
Control Test	0.048	08:38
Air Blank	0.000	08:38
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

P. Murphy
 Operator's Signature

TALLAHASSEE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006926
 10/13/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:30
Control Test	0.078	08:30
Air Blank	0.000	08:31
Control Test	0.079	08:32
Air Blank	0.000	08:32
Control Test	0.079	08:33
Air Blank	0.000	08:33
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

P. Murphy
 Operator's Signature

TALLAHASSEE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006926
 10/13/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:25
Control Test	0.198	08:25
Air Blank	0.000	08:26
Control Test	0.199	08:26
Air Blank	0.000	08:27
Control Test	0.198	08:28
Air Blank	0.000	08:28
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

P. Murphy
 Operator's Signature

TALLAHASSEE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006926
 10/13/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:59
Control Test	0.080	08:59
Air Blank	0.000	08:59
Control Test	0.078	09:00
Air Blank	0.000	09:00
Control Test	0.078	09:01
Air Blank	0.000	09:01
Control Test Stats		
Average	0.0787	
Std Dev	0.0012	
Rel Std Dev(%)	1.4678	

DGS

P. Murphy
 Operator's Signature

TALLAHASSEE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006926
 10/12/2017
 Software: 8100.27

Flow Rate Calibration*****
 1: Rate (Liters/min) = 5
 SQRT(Diff) = 7.000
 2: Rate (Liters/min) = 15
 SQRT(Diff) = 13.000
 3: Rate (Liters/min) = 30
 SQRT(Diff) = 26.191
 Dependent Data Scale Factor = 100000 L/min
 Independent Data Scale Factor = 256
 Rounded Slope = 498
 Rounded Intercept = -297250
 Correlation = 0.99519

SP

10/17/17
 JO