



# INSTRUMENT PROCESSING SHEET

Agency Bay CountyS/N 80-006937

Florida Department of Law Enforcement

Date In 11/20/17DI Completion Date 12/1/17 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>[Signature]</u>	<b>Quality Checks</b> Performed By <u>[Signature]</u>	<b>Flow Calibration</b> Performed By _____
<input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input checked="" type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>169</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.167</u> (.139 - .169) 36 mm <u>.187</u> (.156 - .190) 53 mm <u>.261</u> (.228 - .278) 103 mm <u>.523</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)

**Final Release Date**

**FDLE**

**DEC 08 2017**

**Alcohol Testing Program**

Simulator	Serial #	Lot #/Exp
0.050	<u>G2835</u>	<u>201707D</u> <u>7/25/19</u>
0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>
0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>
0.080 DGS	N/A	<u>AG215202</u> <u>6/1/19</u>

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By [Signature]

Lab Temp °C 21.9  
 External Digital Therm. ID#: 300504  
 34°C +-2 Serial #: G2835  
 34°C +-2 Serial #: DR1279  
 34°C +-2 Serial #: SD1025

**Calibration Adjustment** Performed By \_\_\_\_\_

Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure ID# 28427  
 Gauge 1020 Instrument 1020  
 Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	<u>G2880</u>
Interferent	<u>G2840</u>
0.050	<u>G2835</u>
0.080	<u>DR1279</u>
0.200	<u>SD1025</u>

**Attachments**

Form 41       Flow Calibration  
 Stability Checks     Form 40  
 Calibration Adjustment     Other \_\_\_\_\_  
 Post-Stability Checks

Notes/Suggested Service: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

SP 12/4/17      [Signature] 12/8/17  
 Tech Review / Date      Admin Review / Date

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006937  
12/01/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:21
Control Test	0.048	08:22
Air Blank	0.000	08:23
Control Test	0.049	08:23
Air Blank	0.000	08:24
Control Test	0.048	08:25
Air Blank	0.000	08:25
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

*P Murphy*  
-----  
Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006937  
12/01/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:26
Control Test	0.079	08:27
Air Blank	0.000	08:27
Control Test	0.078	08:28
Air Blank	0.000	08:29
Control Test	0.079	08:29
Air Blank	0.000	08:30
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

*P Murphy*  
-----  
Operator's Signature

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006937  
12/01/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:31
Control Test	0.198	08:32
Air Blank	0.000	08:32
Control Test	0.199	08:33
Air Blank	0.000	08:34
Control Test	0.198	08:34
Air Blank	0.000	08:35
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

*P Murphy*  
-----  
Operator's Signature

*SP*

BAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006937  
12/01/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:39
Control Test	0.080	08:40
Air Blank	0.000	08:40
Control Test	0.081	08:40
Air Blank	0.000	08:41
Control Test	0.080	08:41
Air Blank	0.000	08:42
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

*D6s*

*P Murphy*  
-----  
Operator's Signature

*12/8/17  
JO*