

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 13:20

Date of Inspection: 11/16/2018

Serial Number: 80-000224

Software: 8100.27

JD

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:875088 Exp: 12/12/2020
0.000	0.051	0.082	0.203	0.080
0.000	0.052	0.082	0.202	0.081
0.000	0.051	0.081	0.203	0.080

Number of Simulators Used: 5

Remarks:

WITHIN COMPLIANCE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jeffery A Raker

JEFFERY A RAKER

Signature and Printed Name

11/16/2018
Date