

# Florida Department of Law Enforcement Alcohol Testing Program

1/14/19  
JE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Serial Number: 80-000229

Time of Inspection: 09:57

Date of Inspection: 12/31/2018

Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 201703A<br>Exp: 03/07/2019 | 0.08g/210L Test (g/210L)<br>Lot#: 201808E<br>Exp: 08/22/2020 | 0.20g/210L Test (g/210L)<br>Lot#: 201702B<br>Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: 13518080A5<br>Exp: 08/05/2020 |
|----------------------------|--|--|--|--|
| 0.000                      | 0.048  | 0.078  | 0.194  | 0.078  |
| 0.000                      | 0.049  | 0.079  | 0.196  | 0.078  |
| 0.000                      | 0.049  | 0.078  | 0.196  | 0.078  |

Number of Simulators Used: 3

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC:

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WILLIAM E ROBERSON

Signature and Printed Name

12/31/2018  
Date