

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKALOOSA COUNTY SO
Time of Inspection: 08:10

Date of Inspection: 01/27/2018

Serial Number: 80-000739
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 01603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 20161B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 20702E Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.047	0.078	0.192 / 0.188	
0.000	0.047	0.078	0.188 / 0.193	
0.000	0.047	0.079	0.185 / 0.194	

Number of Simulators Used: 5

Remarks:

20: Control Outside SUSP (C) 5% SIMULATOR. Non-Commercial: SUSPECT. HAD SIMULATOR.

The above instrument complies does not comply with Chapter 119-6, F.S.C.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspection Permit and that I performed this inspection in accordance with the provisions of Chapter 119-6, F.S.C.

D. J. Simon

Signature and Printed Name

01/27/2018 632/129

01/27/2018
Date