

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 14:14

Date of Inspection: 02/21/2018

Serial Number: 80-000752
Software: 8100.27

20

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:644476 Exp: 12/01/2019
/ 0.000	0.049	0.080	0.202	0.081
/ 0.000	0.049	0.080	0.203	0.080
/ 0.000	0.049	0.080	0.203	0.080

Number of Simulators Used: 5

Remarks: AMBIENT FAIL
CLEARED AREA WITH FAN RETESTED INST. PASSED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name JEFFERY A RAKER

02/21/2018
Date