

70-000770 BK

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAMILTON COUNTY SO

Time of Inspection: 17:57

Date of Inspection: 12/30/2018

Serial Number: 80-000770

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 03417080A1 Exp: 03/05/2019
0.000	0.049	0.078	0.193	0.074 / 0.077
0.000	0.049	0.078	0.192	0.075 / 0.076
0.000	0.049	0.079	0.193	0.074 / 0.076

Number of Simulators Used: 4

Remarks:

08: Control Outside Tolerance. *Aired Room out Repeated test.*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

William C Curry
Signature and Printed Name WILLIAM C CURRY

12/30/2018
Date