

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SANTA ROSA COUNTY SO

Time of Inspection: 22:38

Date of Inspection: 12/31/2018

Serial Number: 80-000772

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG805702 Exp: 02/26/2020
0.000	0.052	0.082	0.203	0.079
0.000	0.052	0.082	0.204	0.078
0.000	0.052	0.082	0.204	0.078

Number of Simulators Used: \_\_\_\_\_

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DURON A NELSON  
Signature and Printed Name

12/31/2018  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SANTA ROSA COUNTY SO

Time of Inspection: 19:33

Date of Inspection: 12/31/2018

Serial Number: 80-000772

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK	Yes	No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG809702 Exp: 02/22/2020
0.000	0.049	0.079	0.200	0.074 / 0.074
0.000	0.050	0.079	0.201	0.074 / 0.074
0.000	0.050	0.079	0.200	0.074 / 0.074

Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance, Non-compliance.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DURON A NELSON  
Signature and Printed Name

12/31/2018  
Date



# Santa Rosa Sheriff's Office

## Sheriff Bob Johnson

To: Patrick Murphy, Ph.D.  
Department Inspector, Alcohol Testing Program  
Florida Department of Law Enforcement

From: Sgt. Duron A. Nelson  
Alcohol Testing Program, Agency Inspector

Date: July 30, 2018

Ref: Intoxilyzer 8000  
Serial Number: 80-000772

Dr. Murphy,

Please find the aforementioned breath test instrument accompanying this correspondence, delivered to you for the purpose of the required annual Department Inspection.

Also enclosed, please find the most recent agency inspection performed on 7/30/2018.

V/r

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SANTA ROSA COUNTY SO  
Time of Inspection: 14:28

Date of Inspection: 07/30/2018

Serial Number: 80-000772  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG629801 Exp: 10/24/2018
0.000	0.048	0.077	0.198	0.082
0.000	0.048	0.078	0.199	0.082
0.000	0.049	0.079	0.199	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DURON A NELSON

Signature and Printed Name

07/30/2018  
Date