

78848

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

MAR 07 2018

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: ST LUCIE COUNTY SO  
Time of Inspection: 08:38

Date of Inspection: 02/27/2018

Serial Number: 80-000788  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
RFI / 0.000	0.048	0.081	0.201	0.082
/ 0.000	0.049	0.082	0.201	0.082
/ 0.000	0.049	0.080	0.201	0.082

Number of Simulators Used: 5


Remarks:

00: RFI Detect.

There was a cell phone in the room at the time of the test.  
The cell phone was removed from the room.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_  
Signature and Printed Name

DAVID A CAGLIONI

02/27/2018  
Date