

Florida Department of Law Enforcement Alcohol Testing Program

②

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAMILTON COUNTY SO
Time of Inspection: 01:14

Date of Inspection: 07/30/2018

Serial Number: 80-000797
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019 | 0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03417080A1 Exp: 03/05/2019 |
|----------------------------|---|---|---|---|
| 0.000 | 0.047 | 0.076 | 0.196 | 0.082 |
| 0.000 | 0.047 | 0.077 | 0.198 | 0.081 |
| 0.000 | 0.047 | 0.078 | 0.198 | 0.081 |

Number of Simulators Used: 4

Remarks:
SENDING TO FDLE FOR ANNUAL INSPECTION

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WILLIAM C CURRY
Signature and Printed Name

07/30/2018
Date