

80-000808 BK

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 17:43

Date of Inspection: 12/18/2018

Serial Number: 80-000808
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03417080A1 Exp: 03/05/2019
0.000	0.050	0.081	0.202	0.082
0.000	0.050	0.081	0.202	0.081
0.000	0.051	0.081	0.202	0.081

Number of Simulators Used: 5

Remarks:
AI JDAIL #254236 PRESENT ENTIRE INSPECTION 12-18-18

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name ROGER G SKIPPER

12/18/2018
Date

Reviewed
12-18-18
M.S. 2.1.1. B...
4/18