

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 16 2018

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DIXIE COUNTY SO
Time of Inspection: 09:47

Date of Inspection: 01/11/2018

Serial Number: 80-000809
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A3 Exp: 08/05/2019
0.000	0.045	0.074 / 0.077	0.198	0.080
0.000	0.045	0.075 / 0.077	0.199	0.080
0.000	0.046	0.075 / 0.077	0.200	0.080

Number of Simulators Used: 4

Remarks:

08: Control Outside Tolerance AIR LEAK.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MATTHEW MANDOLA

Signature and Printed Name

01/11/2018
Date