

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED  
MAY 05 2018  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO  
Time of Inspection: 10:26

Date of Inspection: 04/26/2018

Serial Number: 80-000815  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/17/2019 04	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:20016080A3 Exp: 09/05/2018
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.081	0.203	0.081

Number of Simulators Used: 5

Remarks: The expiration date of 03-17-19 was entered incorrectly for the 0.05g. Correct date is 03-07-19. 004 612

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GARY B FORINASH

Signature and Printed Name

04/26/2018  
Date