

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CITRUS COUNTY SO  
Time of Inspection: 10:05

Date of Inspection: 06/29/2018

Serial Number: 80-000817  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 26316080A2 Exp: 11/05/2018
0.000	0.050	0.079	0.192	0.076
0.000	0.049	0.080	0.197	0.077
0.000	0.050	0.079	0.199	0.076

Number of Simulators Used: 3

Remarks:

**RECEIVED**

JUN 29 2018

**FDLE  
Alcohol Testing Program**

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WILLIAM E. ROBERSON

Signature and Printed Name

06/29/2018  
Date