

831 SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO  
Time of Inspection: 14:32

Date of Inspection: 03/15/2018

Serial Number: 80-000831  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 06616080A2 Exp: 04/05/2018
0.000	0.050	0.078	0.198	0.077
0.000	0.050	0.079	0.198	0.077
0.000	0.051	0.079	0.199	0.077

Number of Simulators Used: 5

Remarks:

**RECEIVED**  
MAR 20 2018  
FDLE  
Alcohol Testing Program

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D/S C. [Signature] 5/14/1602 CHRISTOPHER L THOMAS  
Signature and Printed Name

03/15/2018  
Date