RECEIVED
NOV 16 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 20:19 Date of Inspection: 11/15/2018 Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.080	0.198	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

1437 0502 BINDER / PITCHER
Signature and Printed Name

11/15/2018
Date

FDLE/ATP Form 40 – March 2004

80-000836

RECEIVED
OCT 19 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Serial Number: 80-000836
Time of Inspection: 20:50
Date of Inspection: 10/18/2018
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.080	0.200	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

2502 BINDER / PITCHER MTD 1437

Signature and Printed Name

10/18/2018
Date

FDLE/ATP Form 40 – March 2004

80-000836 DMB

RECEIVED
SEP 06 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 12:52 Date of Inspection: 09/04/2018

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.048 / 0.049	0.079	0.197	0.078
0.000	0.049 / 0.049	0.079	0.198	0.078
0.000	0.000 / 0.048	0.079	0.198	0.078

Number of Simulators Used: 5

Remarks:
05: Control Outside Tolerance DISCONNECTED ACCIDENTALLY . ACCIDENTALLY DISCONNECTED

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MLP 1437 2502 PITCHER / BINDER
Signature and Printed Name

09/04/2018
Date

FDLE/ATP Form 40 – March 2004

80-000836 2018

RECEIVED
AUG 16 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 23:27 Date of Inspection: 08/15/2018

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.200	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MID 1437 2502 PITCHER / BINDER
Signature and Printed Name

08/15/2018
Date

FDLE/ATP Form 40 – March 2004

80-000836

RECEIVED
JUL 17 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 23:43 Date of Inspection: 07/06/2018

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.050	0.079	0.199	0.000 / 0.080
0.000	0.050	0.080	0.200	0.000 / 0.079
0.000	0.051	0.080	0.201	0.080 / 0.079

Number of Simulators Used: 5

Remarks:
RECONNECTED DRY GAS HOSE.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MID 1437 2502 PITCHER / BINDER
Signature and Printed Name

07/06/2018
Date

FDLE/ATP Form 40 – March 2004

80-000836

RECEIVED
JUN 11 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 22:18 Date of Inspection: 06/07/2018

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.080	0.199	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DS 4 TC 2502 1437 PITCHER / BINDER
Signature and Printed Name

06/07/2018
Date

80-000836 GMB

RECEIVED
MAY 04 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 17:43 Date of Inspection: 05/03/2018

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.079	0.200	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

F. P 2502 BINDER / PITCHER
Signature and Printed Name

05/03/2018
Date

80-000836 [illegible]

Florida Department of Law Enforcement
Alcohol Testing Program

RECEIVED
JUL 09 2018
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 22:10 Date of Inspection: 04/07/2018
Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:
CLEANED BREATH TEST TUBE . Non-compliance: AMB FAIL. BREATH TEST TUBE HAD WATER DROPS .

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

950 CHESSHER / PITCHER M10 0437
Signature and Printed Name

04/07/2018
Date

FDLE/ATP Form 40 – March 2004

80-000836

RECEIVED
MAR 12 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 18:29 Date of Inspection: 03/09/2018

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.000 / 0.050	0.080	0.200	0.080
0.000	0.000 / 0.050	0.080	0.200	0.080
0.000	0.000 / 0.050	0.080	0.200	0.080

Number of Simulators Used: 5

Remarks:
05:SIM NOT CONNECTED. CONNECTED AND REPEATED

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name PITCHER / CHESSHER

03/09/2018
Date

FDLE/ATP Form 40 – March 2004

80-000836

RECEIVED
FEB 05 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 18:54 Date of Inspection: 02/01/2018

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Pch CL 950

CHESSHER / PITCHER 937
Signature and Printed Name

02/01/2018
Date

FDLE/ATP Form 40 – March 2004

80-000836

RECEIVED
JAN 16 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 18:05 Date of Inspection: 01/04/2018

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.050	0.080	0.199	0.081
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Cpl Jonathan Chessher 950 CHESSHER / PITCHER MID 1437
Signature and Printed Name

01/04/2018
Date

FDLE/ATP Form 40 – March 2004