

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
AUG 15 2018
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NORTH MIAMI BEACH PD
Time of Inspection: 00:19

Date of Inspection: 07/31/2018

Serial Number: 80-000886
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12317080A3 Exp: 07/05/2023 07/05/2019
0.000	0.050	0.078	0.197	0.078
0.000	0.048	0.078	0.194	0.078
0.000	0.049	0.077	0.197	0.078

Number of Simulators Used: 5

Remarks:

AMENDED
08/15/18

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Cora L Mann

Signature and Printed Name

CORA L MANN

07/31/2018
Date