

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAR 08 2018

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 10:53

Date of Inspection: 03/08/2018

Serial Number: 80-000941
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611H Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1981700CA Exp: 09/05/2019
/ 0.000	0.050	0.080	0.198	0.077
/ 0.000	0.050	0.081	0.198	0.077
/ 0.000	0.050	0.081	0.198	0.077

Number of Simulators Used: 5

Remarks: REMOVED AT SPEC OPS RFI on 0.000 g/l

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Handwritten Signature]

Signature and Printed Name

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 12:57

Date of Inspection: 02/06/2018

Serial Number: 80-000941
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) <i>201604C</i> Lot#: 03417080A2 Exp: 03/05/2019 <i>04/05/2018</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 03417080A2 Exp: 03/05/2019
0.000	0.050	0.080	0.000 / 0.197	0.080
0.000	0.050	0.080	0.000 / 0.199	0.080
0.000	0.050	0.080	0.000 / 0.198	0.080


Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance WRONG SOLUTION.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


JOHN E NEIDERHISER
 Signature and Printed Name

02/06/2018
 Date