

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 10:00

Date of Inspection: 05/01/2018

Serial Number: 80-000942
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 19817080A1 Exp: 09/05/2019
0.000	0.050	0.081	0.203	0.078
0.000	0.051	0.082	0.204	0.079
0.000	0.051	0.081	0.204	0.078

Number of Simulators Used: 5

Remarks:

OBSERVED BY NEIDENHISER TESTED AT SPEC OPS



The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Charles E. Baird #202 CHARLES E. BAIRD
Signature and Printed Name

05/01/2018
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 09:09

Date of Inspection: 05/01/2018

Serial Number: 80-000942
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.044 / 0.043			
0.000	0.044 / 0.044			
0.000	0.045 / 0.042			

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: BAD SEAL ON .05 SIMULATOR.

Contacted Dept. Inspector David Reyes-Rivera 5/1/18 0909hrs. advised of situation and the corrective actions taken (bad simulator gasket). Was approved to re-inspect.

2nd inspection completed - test came back with-in range. Jm

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Charles E. Baird 4702

Signature and Printed Name

CHARLES E. BAIRD

05/01/2018
Date