

947  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GLADES COUNTY S.O.  
Time of Inspection: 12:19

Date of Inspection: 02/28/2018

Serial Number: 80-000947  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048	0.070 / 0.071		
0.000	0.049	/ 0.071		
0.000	0.049	/ 0.071		

Number of Simulators Used: 3

Remarks:

08: Improper Sample, Control Outside Tolerance. Non-compliance:

RECEIVED  
JUN 26 2018  
FDLE  
Alcohol Testing Program

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JOHN P GADSON

Signature and Printed Name

02/28/2018  
Date