

*9/17/18*

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY SO  
Time of Inspection: 10:20

Date of Inspection: 09/17/2018

Serial Number: 80-000964  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 201703A<br>Exp: 03/07/2019 | 0.08g/210L Test (g/210L)<br>Lot#: 201709A<br>Exp: 09/12/2017 <i>2019</i> | 0.20g/210L Test (g/210L)<br>Lot#: 201802G<br>Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: 13917080A3<br>Exp: 07/05/2019 |
|----------------------------|--|--|--|--|
| 0.000                      | 0.050  | 0.078  | 0.201  | 0.079  |
| 0.000                      | 0.050  | 0.080  | 0.200  | 0.079  |
| 0.000                      | 0.050  | 0.079  | 0.200  | 0.080  |

Number of Simulators Used: 5

Remarks:

Expiration date of the .08 alcohol Reference Solution used (Lot #201709A) is 09/12/2019, not 09/12/2017 as listed above, *Km*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Kelly M Melville*

Signature and Printed Name KELLY M MELVILLE

09/17/2018  
Date